

## Summary statistics - quantitative results

(Groups: measurement principle)

Filter: Slovakia, minimal size of groups n = 5

## EQA round: ALB2/15 - Albumin in Urine

Dead line: 09.10.2015

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from participants	Nsuc = number of successful participants
	U <sub>AV</sub> = expanded uncertainty of assigned value (k = 2)	Srel = success (relative)

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(369) Albumin</b>					30							30	29	97%
Samples and groups	[mg/L]													
<b>Sample A</b>		172	13	7,6	30	CVP	174	2,1	21%	137,4	210,6	30	29	97%
(1) Immunochemical met.		172	13	7,6	30	0						30		
<b>Sample B</b>		28,1	2,3	8,1	30	CVP	28,7	0,43	30%	20,09	37,31	30	30	100%
(1) Immunochemical met.		28,1	2,3	8,1	30	0						30		
<b>(470) ACR</b>					23							23	21	91%
Samples and groups	[g/mol]													
<b>Sample A</b>		10,1	1,2	12	23	CVP	10,5	0,17	21%	8,295	12,71	23	21	91%
(1) Calculation		10,1	1,2	12	23	0						23		
<b>Sample B</b>		3,92	0,51	13	23	CVP	4,01	0,077	30%	2,807	5,213	23	23	100%
(1) Calculation		3,92	0,51	13	23	0						23		

st\_kn\_p

End of report

Printed: 20.10.2015