

EQA round: CC2/17 - Cystatin C

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(106) Cystatin C					68							68	65	96%
Samples and groups	[mg/L]													
Sample A		0,735	0,068	9,3	68	CVP	0,735	0,020	18%	0,602	0,868	68	65	96%
All results		0,735	0,068	9,3	68	0						68		
Sample B		3,65	0,20	5,5	68	CVP	3,65	0,060	18%	2,99	4,31	68	68	100%
All results		3,65	0,20	5,5	68	0						68		

st_kn_p

End of report

Printed: 18.10.2017