

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]						Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(690) WBC					22							22	20	91%
Samples and groups	[·10(exp9)/L]													
Sample A		23,6	0,87	3,7	22	CVP	23,6	0,46	15%	20	27,2	22	22	100%
All results		23,6	0,87	3,7	22	0						22		
Sample B		7,95	0,40	5,0	22	CVP	7,95	0,21	15%	6,75	9,15	22	20	91%
All results		7,95	0,40	5,0	22	0						22		
(691) CD 34+ (relative count)					22							22	18	82%
Samples and groups	[%]													
Sample A		1,03	0,053	5,2	22	CVP	1,03	0,028	25%	0,772	1,29	22	21	95%
All results		1,03	0,053	5,2	22	0						22		
Sample B		0,033	0,011	34	22	CVP	0,033	0,0059 ± 0,025		0,008	0,058	22	19	86%
All results		0,033	0,011	34	22	0						22		
(692) CD 34+ (absolute count)					22							22	19	86%
Samples and groups	[count/μL]													
Sample A		243	19	7,7	22	CVP	243	9,8	25%	182	304	22	21	95%
All results		243	19	7,7	22	0						22		
Sample B		2,84	0,90	32	22	CVP	2,84	0,48 ± 1,25		1,59	4,09	22	19	86%
All results		2,84	0,90	32	22	1						22		

st_kn_p

End of report

Printed: 28.10.2017