

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U_{AV} = expanded uncertainty of the assigned value ($k = 2$)	

Test	[unit]	Comparability						AV	U_{AV}	Dmax	LL	UL	N _{eva}	N _{suc}	S _{rel}
		RoM	SD	CV [%]	N _{tot}	N _{out}									
(130) WBC													16	16	100%
Samples and groups	f.10(exp9)/L]														
Sample A		5,46	0,27	5,0	16		CVP	5,58	0,046	15%	4,74	6,42	16	16	100%
All results		5,46	0,27	5,0	16	0							16		
Sample B		4,42	0,21	4,7	16		CVP	4,45	0,038	15%	3,78	5,12	16	16	100%
All results		4,42	0,21	4,7	16	0							16		
(160) Neutrophiles													16	15	94%
Samples and groups	f.10(exp9)/L]														
Sample A		3,43	0,25	7,3	16		CVP	3,48	0,038	25%	2,61	4,35	16	15	94%
All results		3,43	0,25	7,3	16	0							16		
Sample B		2,56	0,18	7,2	16		CVP	2,56	0,026	25%	1,92	3,2	16	16	100%
All results		2,56	0,18	7,2	16	0							16		
(161) Lymphocytes													16	16	100%
Samples and groups	f.10(exp9)/L]														
Sample A		1,47	0,18	13	16		CVP	1,51	0,023	25%	1,13	1,89	16	16	100%
All results		1,47	0,18	13	16	0							16		
Sample B		1,31	0,13	9,8	16		CVP	1,33	0,020	25%	0,997	1,67	16	16	100%
All results		1,31	0,13	9,8	16	0							16		
(162) Monocytes													16	14	88%
Samples and groups	f.10(exp9)/L]														
Sample A		0,374	0,065	17	16		CVP	0,366	0,013	50%	0,183	0,549	16	15	94%
All results		0,374	0,065	17	16	0							16		
Sample B		0,318	0,061	19	16		CVP	0,313	0,011	50%	0,156	0,47	16	14	88%
All results		0,318	0,061	19	16	0							16		
(163) Eosinophiles													16	16	100%
Samples and groups	f.10(exp9)/L]														
Sample A		0,209	0,032	15	16		CVP	0,208	,0035	50%	0,104	0,312	16	16	100%
All results		0,209	0,032	15	16	0							16		
Sample B		0,224	0,025	11	16		CVP	0,219	,0046	50%	0,109	0,329	16	16	100%
All results		0,224	0,025	11	16	0							16		
(164) Basophiles													16	16	100%
Samples and groups	f.10(exp9)/L]														
Sample A		0,015	0,012	83	16		CVP	0,031	,0042	asym.	0	0,167	16	16	100%
All results		0,015	0,012	83	16	0							16		
Sample B		0,014	0,009	69	16		CVP	0,025	,0032	asym.	0	0,134	16	16	100%
All results		0,014	0,009	69	16	0							16		