

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	$U_{AV}$ = expanded uncertainty of the assigned value ( $k = 2$ )	

Test	[unit]						Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	$U_{AV}$	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(130) WBC</b>					29							29	29	100%
Samples and groups	f.10(exp9)/L													
<b>Sample A</b>		5,46	0,23	4,3	29	CVP	5,53	0,033	15%	4,7	6,36	29	29	100%
All results		5,46	0,23	4,3	29	0						29	29	100%
<b>Sample B</b>		4,42	0,19	4,3	29	CVP	4,48	0,025	15%	3,8	5,16	29	29	100%
All results		4,42	0,19	4,3	29	0						29	29	100%
<b>(160) Neutrophiles</b>					29							29	28	97%
Samples and groups	f.10(exp9)/L													
<b>Sample A</b>		3,49	0,17	5,0	29	CVP	3,49	0,028	25%	2,61	4,37	29	28	97%
All results		3,49	0,17	5,0	29	1						29	28	97%
<b>Sample B</b>		2,64	0,14	5,4	29	CVP	2,62	0,019	25%	1,96	3,28	29	28	97%
All results		2,64	0,14	5,4	29	1						29	28	97%
<b>(161) Lymphocytes</b>					29							29	24	83%
Samples and groups	f.10(exp9)/L													
<b>Sample A</b>		1,36	0,16	12	29	CVP	1,39	0,015	25%	1,04	1,74	29	24	83%
All results		1,36	0,16	12	29	1						29	24	83%
<b>Sample B</b>		1,40	0,10	7,3	29	CVP	1,43	0,012	25%	1,07	1,79	29	27	93%
All results		1,40	0,10	7,3	29	1						29	27	93%
<b>(162) Monocytes</b>					29							29	28	97%
Samples and groups	f.10(exp9)/L													
<b>Sample A</b>		0,426	0,079	18	29	CVP	0,445	,0099	50%	0,222	0,668	29	28	97%
All results		0,426	0,079	18	29	0						29	28	97%
<b>Sample B</b>		0,282	0,066	23	29	CVP	0,305	,0075	50%	0,152	0,458	29	28	97%
All results		0,282	0,066	23	29	0						29	28	97%
<b>(163) Eosinophiles</b>					29							29	21	72%
Samples and groups	f.10(exp9)/L													
<b>Sample A</b>		0,163	0,051	32	29	CVP	0,176	,0044	50%	0,088	0,264	29	23	79%
All results		0,163	0,051	32	29	0						29	23	79%
<b>Sample B</b>		0,094	0,023	24	29	CVP	0,098	,0022	asym.	0	0,147	29	26	90%
All results		0,094	0,023	24	29	0						29	26	90%
<b>(164) Basophiles</b>					29							29	29	100%
Samples and groups	f.10(exp9)/L													
<b>Sample A</b>		0,022	0,014	63	29	CVP	0,03	,0027	asym.	0	0,166	29	29	100%
All results		0,022	0,014	63	29	0						29	29	100%
<b>Sample B</b>		0,008	0,006	80	29	CVP	0,013	,0017	asym.	0	0,134	29	29	100%
All results		0,008	0,006	80	29	0						29	29	100%