

Key: &gt;&gt;&gt; ... marks correct (expected) result

&gt; ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(261) Glucose</b>				
A	All results	>>> 10	11,1 - 28 mmol/L	45 %
		>>> 12	> 28 mmol/L	55 %
	Total per sample: n = 22	Success: 100 %		
B	All results	>>> 17	Negative, normal, <= 2,8 mmol/L	77 %
		> 5	2,81 - 11 mmol/L	23 %
	Total per sample: n = 22	Success: 100 %		
Total success: 100 %				

**(263) Ketone bodies**

A	All results	>>> 12	1+, 2+	55 %
		>>> 10	3+, 4+	45 %
	Total per sample: n = 22	Success: 100 %		
B	All results	>>> 22	Negative, <= 0,5 mmol/L	100 %
	Total per sample: n = 22	Success: 100 %		
Total success: 100 %				

**(264) Total protein**

A	All results	>>> 10	0,51 - 2,5 g/L	45 %
		>>> 12	> 2,5 g/L	55 %
	Total per sample: n = 22	Success: 100 %		
B	All results	>>> 22	Negative, < 0,3 g/L	100 %
	Total per sample: n = 22	Success: 100 %		
Total success: 100 %				

**(266) Urobilinogen**

A	All results	>>> 5	1+, 2+	23 %
		>>> 17	3+, 4+	77 %
	Total per sample: n = 22	Success: 100 %		
B	All results	>>> 22	Negative, < 17 µmol/L	100 %
	Total per sample: n = 22	Success: 100 %		
Total success: 100 %				

**(268) Bilirubin**

A	All results	> 4	1+, 2+	19 %
		>>> 17	3+, 4+	81 %
	Total per sample: n = 21	Success: 100 %		
B	All results	>>> 21	Negative	100 %
	Total per sample: n = 21	Success: 100 %		
Total success: 100 %				

**(270) WBC**

A	All results (without individual groups)	>>> 6	51 - 250 leu/µL	29 %
		>>> 12	> 250 leu/µL	57 %
	(223) Iris iChem Velocity	± 2	Negative	9,5 %
		± 1	51 - 250 leu/µL	4,8 %
	Total per sample: n = 21	Success (only evaluated results): 100 %		
B	All results	>>> 21	Negative	100 %
	Total per sample: n = 21	Success: 100 %		
Total success (only evaluated tests): 100 %				

**(272) Blood**

A	All results	1	5 - 12 ery/µL	4,5 %
		>>> 21	> 50 ery/µL	95 %
	Total per sample: n = 22	Success: 95 %		
B	All results	>>> 22	Negative	100 %
	Total per sample: n = 22	Success: 100 %		
Total success: 95 %				

**(273) pH**

A	All results	>>> 5	6,5	23 %
		>>> 17	7,0	77 %
	Total per sample: n = 22	Success: 100 %		
B	All results	> 2	5,0	9,1 %
		>>> 1	6,0	4,5 %
		>>> 17	6,5	77 %
		> 2	7,0	9,1 %
	Total per sample: n = 22	Success: 100 %		
Total success: 100 %				

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(274) Nitrite</b>				
A	All results	>>> 21	Positive	100 %
----- Total per sample: n = 21		----- Success: 100 %	-----	
B	All results	>>> 21	Negative	100 %
----- Total per sample: n = 21		----- Success: 100 %	-----	
Total success: 100 %				

**(275) hCG**

A	All results	>>> 5	Positive	100 %
----- Total per sample: n = 5		----- Success: 100 %	-----	
B	All results	>>> 5	Negative	100 %
----- Total per sample: n = 5		----- Success: 100 %	-----	
Total success: 100 %				

**(276) Specific gravity**

A	All results	± 2	1,005	11 %
		± 4	1,010	22 %
		± 2	1,015	11 %
		± 7	1,020	39 %
		± 2	1,025	11 %
		± 1	1,030	5,6 %
----- Total per sample: n = 18		----- Success (only evaluated results): evaluation-----		
B	All results	> 5	1,005	28 %
		>>> 10	1,010	56 %
		> 3	1,015	17 %
----- Total per sample: n = 18		----- Success: 100 % -----		