

## EQA round: FC2/17 - Calprotectin in Stool

Dead line: 22.09.2017

Key: &gt;&gt;&gt; ... marks correct (expected) result

&gt; ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

## Test

Sample	Group	N	Result	N <sub>rel</sub>
<b>(223) Calprotectin</b>				
A	(49) BioVendor	>>> 2	Positive	3,9 %
	(87) Euroimmun	>>> 1	Positive	2,0 %
	(135) OrgenTec	>>> 10	Positive	20 %
	(164) DiaSorin	>>> 2	Positive	3,9 %
	(166) Thermo	>>> 2	Positive	3,9 %
	(219) Bühlmann	1	Negative	2,0 %
		1	Inconclusive result (borderline)	2,0 %
		>>> 21	Positive	41 %
	(999) another manufacturer	1	Negative	2,0 %
		1	Inconclusive result (borderline)	2,0 %
		>>> 9	Positive	18 %
----- Total per sample: n = 51 ----- Success: 92 % -----				
B	(49) BioVendor	>>> 2	Positive	3,9 %
	(87) Euroimmun	>>> 1	Positive	2,0 %
	(135) OrgenTec	>>> 10	Positive	20 %
	(164) DiaSorin	>>> 2	Positive	3,9 %
	(166) Thermo	>>> 2	Positive	3,9 %
	(219) Bühlmann	2	Inconclusive result (borderline)	3,9 %
		>>> 21	Positive	41 %
	(999) another manufacturer	1	Negative	2,0 %
		2	Inconclusive result (borderline)	3,9 %
		>>> 8	Positive	16 %
----- Total per sample: n = 51 ----- Success: 90 % -----				
Total success: 90 %				