

Summary statistics - quantitative results

(Groups: manufacturer of instrument)

Filter: minimal size of groups n = 5

EQA round: HS2/17 - Haemocoagulation Special

Dead line: 27.10.2017

Test	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability						N _{eva}	N _{suc}	S _{rel}
							AV	U _{AV}	D _{max}	LL	UL				
(486) Protein S					76								76	74	97%
Samples and groups	[%]														
Sample A3		83,1	13	16	76		CVP	83,1	3,7	30%	58,1	109	76	74	97%
(38) IL		88,4	9,1	10	13	0							13		
(63) Sysmex		82,5	13	16	41	0							41		
(94) Stago		96,0	22	23	7	0							7		
(149) Siemens (Dade)		74,0	10	14	13	0							13		
Other					2	0							2		
							1x 12, 1x 51								
(487) ProC Global (norm. ratio)					37								37	37	100%
Samples and groups	[-]														
Sample A3		0,443	0,029	6,5	37		CVP	0,443	0,012	25%	0,332	0,554	37	37	100%
(63) Sysmex		0,443	0,033	7,4	24	0							24		
(149) Siemens (Dade)		0,440	0,015	3,4	7	0							7		
Other					6	0							6		
							1x 27, 1x 38, 4x 94								
Set of samples 4															
(491) Factor II					33								33	33	100%
Samples and groups	[%]														
Sample A4		86,0	7,0	8,2	33		CVP	86	3,0	25%	64,5	108	33	33	100%
(63) Sysmex		86,3	7,0	8,1	18	0							18		
(94) Stago		83,5	6,7	8,0	5	0							5		
(149) Siemens (Dade)		91,8	6,2	6,8	6	0							6		
Other					4	0							4		
							1x 12, 3x 38								
(492) Factor V					36								36	36	100%
Samples and groups	[%]														
Sample A4		107	7,6	7,2	36		CVP	107	3,1	25%	80,2	134	36	36	100%
(63) Sysmex		107	8,5	7,9	21	0							21		
(94) Stago		103	4,4	4,3	5	0							5		
(149) Siemens (Dade)		107	2,4	2,3	6	0							6		
Other					4	0							4		
							1x 12, 3x 38								
(493) Factor VII					38								38	34	89%
Samples and groups	[%]														
Sample A4		11,7	2,1	18	38		CVP	11,7	0,82	36%	7,48	16	38	34	89%
(63) Sysmex		11,2	1,7	15	23	0							23		
(94) Stago		12,0	3,0	25	5	0							5		
(149) Siemens (Dade)		11,2	0,52	4,7	6	0							6		
Other					4	0							4		
							1x 12, 3x 38								
(494) Factor X					30								30	30	100%
Samples and groups	[%]														
Sample A4		87,0	6,3	7,3	30		CVP	87	2,8	25%	65,2	109	30	30	100%
(63) Sysmex		86,3	5,1	5,9	15	0							15		
(94) Stago		85,5	2,2	2,6	5	0							5		
(149) Siemens (Dade)		87,6	11	13	6	0							6		
Other					4	0							4		
							1x 12, 3x 38								

st_kn_p

End of report

Printed: 09.11.2017