

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	$U_{AV}$ = expanded uncertainty of the assigned value ( $k = 2$ )	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	$U_{AV}$	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(253) CD 3+ T lymphocytes</b>					31							31	29	94%
Samples and groups	[%]													
<b>Sample A</b>		53,6	1,6	2,9	31	CVP	53,5	0,42	7%	49,7	57,3	31	30	97%
All results		53,6	1,6	2,9	31	0						31		
<b>Sample B</b>		71,6	1,5	2,1	31	CVP	71,3	0,44	7%	66,3	76,3	31	30	97%
All results		71,6	1,5	2,1	31	0						31		
<b>(254) CD 4+ T lymphocytes</b>					31							31	30	97%
Samples and groups	[%]													
<b>Sample A</b>		38,2	1,5	4,0	31	CVP	38,2	0,42	10%	34,3	42,1	31	31	100%
All results		38,2	1,5	4,0	31	0						31		
<b>Sample B</b>		47,1	1,5	3,1	31	CVP	46,7	0,38	10%	42	51,4	31	30	97%
All results		47,1	1,5	3,1	31	0						31		
<b>(255) CD 8+ T lymphocytes</b>					31							31	27	87%
Samples and groups	[%]													
<b>Sample A</b>		13,6	0,88	6,5	31	CVP	13,7	0,23	17%	11,3	16,1	31	29	94%
All results		13,6	0,88	6,5	31	0						31		
<b>Sample B</b>		19,8	1,4	7,2	31	CVP	20	0,36	17%	16,6	23,4	31	29	94%
All results		19,8	1,4	7,2	31	0						31		
<b>(256) B lymphocytes</b>					31							31	28	90%
Samples and groups	[%]													
<b>Sample A</b>		2,88	0,41	14	31	CVP	2,95	0,11	30%	2,06	3,84	31	28	90%
All results		2,88	0,41	14	31	0						31		
<b>Sample B</b>		12,7	1,2	9,3	31	CVP	12,9	0,29	30%	9,03	16,8	31	31	100%
All results		12,7	1,2	9,3	31	0						31		
<b>(257) NK cells</b>					31							31	31	100%
Samples and groups	[%]													
<b>Sample A</b>		41,9	2,1	5,1	31	CVP	42,3	0,53	26%	31,3	53,3	31	31	100%
All results		41,9	2,1	5,1	31	0						31		
<b>Sample B</b>		13,5	1,6	12	31	CVP	13,7	0,37	26%	10,1	17,3	31	31	100%
All results		13,5	1,6	12	31	0						31		