

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
Set of samples 1														
(536) Prothrombin test (CoaguChek)					130							130	127	98%
Samples and groups														
Sample A1		3,28	0,17	5,2	130	CVP	3,28	0,037	20%	2,62	3,94	130	130	100%
All results		3,28	0,17	5,2	130	0						130		
Sample B1		3,02	0,21	7,1	130	CVP	3,02	0,046	20%	2,41	3,63	130	127	98%
All results		3,02	0,21	7,1	130	0						130		
st_kn_p						End of report								Printed: 26.06.2017