

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability										
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}

Set of samples 1

(536) Prothrombin test (CoaguChek)		163					163			162		99%			
Samples and groups		[-]													
Sample A1		1,44	0,059	4,1	163		CVP	1,44	0,011	20%	1,15	1,73	163	162	99%
All results		1,44	0,059	4,1	163	0							163		
Sample B1		1,44	0,057	3,9	163		CVP	1,44	0,011	20%	1,15	1,73	163	163	100%
All results		1,44	0,057	3,9	163	0							163		

st_kn_p

End of report

Printed: 30.11.2017