

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]						Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(150) WBC</b>					49							49	48	98%
Samples and groups	f.10(exp9)/L													
<b>Sample A</b>		10,5	0,48	4,5	49	CVP	10,6	0,079	15%	9,01	12,2	49	48	98%
All results		10,5	0,48	4,5	49	0						49		
<b>Sample B</b>		3,76	0,20	5,2	49	CVP	3,76	0,033	18%	3,08	4,44	49	49	100%
All results		3,76	0,20	5,2	49	0						49		
<b>(151) RBC</b>					49							49	49	100%
Samples and groups	f.10(exp12)/L													
<b>Sample A</b>		3,50	0,055	1,6	49	CVP	3,51	0,012	7%	3,26	3,76	49	49	100%
All results		3,50	0,055	1,6	49	0						49		
<b>Sample B</b>		6,39	0,098	1,5	49	CVP	6,42	0,021	7%	5,97	6,87	49	49	100%
All results		6,39	0,098	1,5	49	0						49		
<b>(152) Haemoglobin</b>					49							49	48	98%
Samples and groups	[g/L]													
<b>Sample A</b>		90,9	1,5	1,6	49	CVP	90,7	0,28	6%	85,2	96,2	49	48	98%
All results		90,9	1,5	1,6	49	0						49		
<b>Sample B</b>		171	3,2	1,9	49	CVP	169	0,55	6%	158	180	49	49	100%
All results		171	3,2	1,9	49	0						49		
<b>(153) HCT</b>					49							49	47	96%
Samples and groups	[-]													
<b>Sample A</b>		0,252	0,011	4,5	49							49	48	98%
(1) Automate		0,253	0,011	4,2	46	0	CVPG	0,261	0,024	10%	0,234	0,288	46	
Other					3	0							3	
														3x2
<b>Sample B</b>		0,475	0,019	4,0	49							49	48	98%
(1) Automate		0,476	0,018	3,8	46	0	CVPG	0,492	0,039	10%	0,442	0,542	46	
Other					3	0							3	
														3x2
<b>(154) MCV</b>					49							49	49	100%
Samples and groups	[fL]													
<b>Sample A</b>		72,1	2,7	3,7	49							49	49	100%
(1) Automate		72,3	2,4	3,4	46	0	CVPG	74,2	0,59	10%	66,7	81,7	46	
Other					3	0							3	
														3x2
<b>Sample B</b>		74,4	2,7	3,7	49							49	49	100%
(1) Automate		74,7	2,6	3,5	46	0	CVPG	76,7	0,60	10%	69	84,4	46	
Other					3	0							3	
														3x2
<b>(155) Platelets</b>					49							49	49	100%
Samples and groups	f.10(exp9)/L													
<b>Sample A</b>		509	36	7,2	49							49	49	100%
All results (without individual groups)		510	36	7,0	48	0	CVP	504	6,6	20%	403	605	48	
Other					1	0							1	
														1x3
<b>Sample B</b>		179	18	10	49							49	49	100%
All results (without individual groups)		180	18	9,9	48	0	CVP	178	3,3	27%	129	227	48	
Other					1	0							1	
														1x3
<b>(156) RDW</b>					40							40	34	85%
Samples and groups	[%]													
<b>Sample A</b>		13,8	2,4	17	40							40	36	90%
(749) Sysmex XE, XS, XT series		15,1	0,150	9,9	17	0	CVPG	15,2	0,082	10%	13,6	16,8	17	
(755) Sysmex KX series		10,8	0,65	6,0	12	1	CVPG	10,7	0,36	10%	9,63	11,8	12	
Other					11	0							11	
														1x737, 2x765, 2x766, 2x768, 2x772, 2x773
<b>Sample B</b>		16,6	3,0	18	40							40	34	85%
(749) Sysmex XE, XS, XT series		18,5	1,2	6,6	17	0	CVPG	18,4	0,47	10%	16,5	20,3	17	
(755) Sysmex KX series		13,5	0,76	5,6	12	1	CVPG	13,3	0,42	10%	11,9	14,7	12	
Other					11	0							11	
														1x737, 2x765, 2x766, 2x768, 2x772, 2x773
<b>(157) MPV</b>					43							42	41	98%
Samples and groups	[fL]													
<b>Sample A</b>		9,63	0,46	4,8	43							42	41	98%
(1) Automate; (63) Sysmex		9,65	0,37	3,8	34	0	CVPG	9,85	0,099	18%	8,07	11,7	34	

Test	[unit]	RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	Comparability							
							AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(157) MPV</b>	[fL]				43							42	41	98%
Samples and groups														
<b>Sample A</b>		9,63	0,46	4,8	43							42	41	98%
Other					9	0						8		
							2x 1/12, 1x 1/176, 3x 1/177, 1x 1/179, 2x 2/1							
<b>Sample B</b>		9,49	0,38	4,0	43							42	42	100%
(1) Automate; (63) Sysmex		9,46	0,38	4,0	34	0	CVPG	9,61	0,079	18%	7,88	11,4	34	
Other					9	0						8		
							2x 1/12, 1x 1/176, 3x 1/177, 1x 1/179, 2x 2/1							
<b>(158) PDW [%]</b>	[%]				3							2	2	100%
Samples and groups														
<b>Sample A</b>		21,3	10	49	3							2	2	100%
Other					3	0						2		
							1x 12, 1x 176, 1x 179							
<b>Sample B</b>		31,0	23	73	3							2	2	100%
Other					3	0						2		
							1x 12, 1x 176, 1x 179							
<b>(165) PDW [fL]</b>	[fL]				32							32	31	97%
Samples and groups														
<b>Sample A</b>		9,39	0,45	4,8	32		CVP	9,35	0,10	15%	7,94	10,8	32	31
All results		9,39	0,45	4,8	32	0						32		
<b>Sample B</b>		10,9	0,87	8,0	32		CVP	10,5	0,16	15%	8,92	12,1	32	32
All results		10,9	0,87	8,0	32	0						32		
<b>(166) PDW [-]</b>	[-]				4							4	4	100%
Samples and groups														
<b>Sample A</b>		15,8	2,2	14	4							4	4	100%
Other					4	0						4		
							1x 1, 3x 177							
<b>Sample B</b>		17,0	2,2	13	4							4	4	100%
Other					4	0						4		
							1x 1, 3x 177							

st\_kn\_p

End of report

Printed: 02.08.2017