

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of groups n = 5

EQA round: LMWH2/17 - Low Molecular Weight Heparin

Dead line: 27.10.2017

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U_{AV} = expanded uncertainty of the assigned value ($k = 2$)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U_{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(103) Low Molecular Weight Heparin					11							11	8	73%
Samples and groups	[IU/mL]													
Sample A		0,802	0,092	12	11	CVP	0,781	0,016	25%	0,585	0,977	11	10	91%
(193) HYPHEN BioMed		0,845	0,067	7,9	7							7		
Other					4							4		
						1x 38, 3x 149								
Sample B		0,270	0,056	21	11	CVP	0,249	,0089	30%	0,174	0,324	11	8	73%
(193) HYPHEN BioMed		0,290	0,059	20	7							7		
Other					4							4		
						1x 38, 3x 149								
st_kn_p						End of report								Printed: 09.11.2017