

Key: >>> ... marks correct (expected) result
> ... marks conditionally correct (acceptable) result
± ... marks the result not evaluated

N ... number of the results
Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(1001) DAT (direct Coombs test)				
1	All results	1	Negative	1,3 %
		>>> 41	Positive (1+)	54 %
		>>> 30	Positive (2+)	39 %
		> 4	Positive (3+)	5,3 %
----- Total per sample: n = 76 ----- Success: 99 % -----				
2	All results	> 27	Positive (2+)	36 %
		>>> 47	Positive (3+)	62 %
		> 2	Positive (4+)	2,6 %
----- Total per sample: n = 76 ----- Success: 100 % -----				
Total success: 99 %				
(1002) Sensibilization specificity anti-IgG				
1	All results	>>> 28	Positive (1+)	70 %
		>>> 11	Positive (2+)	28 %
		> 1	Positive (3+)	2,5 %
----- Total per sample: n = 40 ----- Success: 100 % -----				
2	All results	> 1	Positive (1+)	2,5 %
		> 12	Positive (2+)	30 %
		>>> 26	Positive (3+)	65 %
		> 1	Positive (4+)	2,5 %
----- Total per sample: n = 40 ----- Success: 100 % -----				
Total success: 100 %				
(1003) Sensibilization specificity anti-IgA				
1	All results	>>> 30	Negative	100 %
----- Total per sample: n = 30 ----- Success: 100 % -----				
2	All results	>>> 30	Negative	100 %
----- Total per sample: n = 30 ----- Success: 100 % -----				
Total success: 100 %				
(1004) Sensibilization specificity anti-IgM				
1	All results	>>> 30	Negative	100 %
----- Total per sample: n = 30 ----- Success: 100 % -----				
2	All results	>>> 30	Negative	100 %
----- Total per sample: n = 30 ----- Success: 100 % -----				
Total success: 100 %				
(1005) Sensibilization specificity anti-C3c				
1	All results	>>> 30	Negative	100 %
----- Total per sample: n = 30 ----- Success: 100 % -----				
2	All results	>>> 30	Negative	100 %
----- Total per sample: n = 30 ----- Success: 100 % -----				
Total success: 100 %				
(1006) Sensibilization specificity anti-C3d				
1	All results	>>> 37	Negative	100 %
----- Total per sample: n = 37 ----- Success: 100 % -----				
2	All results	>>> 37	Negative	100 %
----- Total per sample: n = 37 ----- Success: 100 % -----				
Total success: 100 %				
(1008) Col. agglut: Quantif. IgG-DAT				
1	All results	> 5	Not specified	26 %
		>>> 9	10	47 %
		> 5	Evaluation impossible	26 %
----- Total per sample: n = 19 ----- Success: 100 % -----				
2	All results	> 1	10	5,3 %
		>>> 18	30	95 %
----- Total per sample: n = 19 ----- Success: 100 % -----				
Total success: 100 %				
(1009) Col. agglut: Subclass IgG1 1:1				
1	All results	> 2	Not specified	10 %
		>>> 17	Negative	85 %
		1	Positive (1+)	5,0 %
----- Total per sample: n = 20 ----- Success: 95 % -----				
2	All results	>>> 19	Negative	95 %
		1	Positive (2+)	5,0 %
----- Total per sample: n = 20 ----- Success: 95 % -----				
Total success: 95 %				

Test		N	Result	N _{rel}
Sample	Group			
(1010) Col. agglut: Subclass IgG1 1:100				
1	All results	>	2 Not specified	10 %
		>>>	18 Negative	90 %
	Total per sample: n = 20	Success: 100 %		
2	All results	>>>	20 Negative	100 %
	Total per sample: n = 20	Success: 100 %		
Total success: 100 %				
(1011) Col. agglut: Subclass IgG3 1:1				
1	All results	>	2 Not specified	10 %
		>>>	18 Negative	90 %
	Total per sample: n = 20	Success: 100 %		
2	All results	>>>	20 Negative	100 %
	Total per sample: n = 20	Success: 100 %		
Total success: 100 %				
(1012) Col. agglut: Subclass IgG3 1:100				
1	All results	>	2 Not specified	10 %
		>>>	18 Negative	90 %
	Total per sample: n = 20	Success: 100 %		
2	All results	>>>	20 Negative	100 %
	Total per sample: n = 20	Success: 100 %		
Total success: 100 %				
(1014) Tube test: Quantification IgG-DAT				
1	All results		1 titre 8	50 %
			1 Evaluation impossible	50 %
	Total per sample: n = 2	Success: 0,00 %		
2	All results		2 titre 2	100 %
	Total per sample: n = 2	Success: 0,00 %		
Total success: 0,00 %				
(1019) ABO blood group				
1	All results	>>>	48 Blood group: AB	100 %
	Total per sample: n = 48	Success: 100 %		
2	All results	>>>	48 Blood group: A	100 %
	Total per sample: n = 48	Success: 100 %		
Total success: 100 %				
(1020) Rh D antigen				
1	All results	>>>	47 Negative (RhD-)	98 %
			1 Positive (RhD+)	2,1 %
	Total per sample: n = 48	Success: 98 %		
2	All results	>>>	48 Positive (RhD+)	100 %
	Total per sample: n = 48	Success: 100 %		
Total success: 98 %				
(1021) Antigen C				
1	All results	>>>	38 Negative	100 %
	Total per sample: n = 38	Success: 100 %		
2	All results	>>>	38 Positive	100 %
	Total per sample: n = 38	Success: 100 %		
Total success: 100 %				
(1022) Antigen c				
1	All results	>>>	38 Positive	100 %
	Total per sample: n = 38	Success: 100 %		
2	All results	>>>	1 Negative	2,6 %
		>>>	37 Positive	97 %
	Total per sample: n = 38	Success: 97 %		
Total success: 97 %				
(1023) Antigen Cw				
1	All results		1 Not specified	2,9 %
		>>>	34 Negative	97 %
	Total per sample: n = 35	Success: 97 %		
2	All results	>>>	35 Negative	100 %
	Total per sample: n = 35	Success: 100 %		
Total success: 97 %				
(1024) Antigen E				
1	All results	>>>	38 Negative	100 %
	Total per sample: n = 38	Success: 100 %		
2	All results	>>>	1 Not specified	2,6 %
		>>>	36 Negative	95 %
			1 Positive	2,6 %

Test

Sample	Group	N	Result	N _{rel}
Total per sample: n = 38		Success: 95 %		
Total success: 95 %				
(1025) Antigen e				
1	All results	>>> 38	Positive	100 %
Total per sample: n = 38		Success: 100 %		
2	All results	>>> 38	Positive	100 %
Total per sample: n = 38		Success: 100 %		
Total success: 100 %				
(1026) Antigen K				
1	All results	>>> 38	Negative	100 %
Total per sample: n = 38		Success: 100 %		
2	All results	1 Negative >>> 37 Positive		2,6 % 97 %
Total per sample: n = 38		Success: 97 %		
Total success: 97 %				
(1027) Antigen k				
1	All results	>>> 22	Positive	100 %
Total per sample: n = 22		Success: 100 %		
2	All results	1 Negative >>> 21 Positive		4,5 % 95 %
Total per sample: n = 22		Success: 95 %		
Total success: 95 %				
(1028) Antigen M				
1	All results	>>> 24	Positive	100 %
Total per sample: n = 24		Success: 100 %		
2	All results	>>> 24	Positive	100 %
Total per sample: n = 24		Success: 100 %		
Total success: 100 %				
(1029) Antigen N				
1	All results	>>> 22	Positive	100 %
Total per sample: n = 22		Success: 100 %		
2	All results	>>> 22	Negative	100 %
Total per sample: n = 22		Success: 100 %		
Total success: 100 %				
(1030) Antigen S				
1	All results	>>> 21 Negative 1 Positive		95 % 4,5 %
Total per sample: n = 22		Success: 95 %		
2	All results	>>> 22	Positive	100 %
Total per sample: n = 22		Success: 100 %		
Total success: 95 %				
(1031) Antigen s				
1	All results	1 Negative >>> 15 Positive		6,3 % 94 %
Total per sample: n = 16		Success: 94 %		
2	All results	>>> 16	Positive	100 %
Total per sample: n = 16		Success: 100 %		
Total success: 94 %				
(1032) Antigen P1				
1	All results	>>> 21	Positive	100 %
Total per sample: n = 21		Success: 100 %		
2	All results	>>> 21	Positive	100 %
Total per sample: n = 21		Success: 100 %		
Total success: 100 %				
(1033) Antigen Jka				
1	All results	>>> 25	Positive	100 %
Total per sample: n = 25		Success: 100 %		
2	All results	>>> 25	Positive	100 %
Total per sample: n = 25		Success: 100 %		
Total success: 100 %				
(1034) Antigen Jkb				
1	All results	>>> 25	Positive	100 %
Total per sample: n = 25		Success: 100 %		
2	All results	>>> 25	Positive	100 %
Total per sample: n = 25		Success: 100 %		
Total success: 100 %				

Test

Sample	Group	N	Result	N _{rel}
(1035) Antigen Lea				
1	All results	>>> 20	Negative	100 %
----- Total per sample: n = 20		----- Success: 100 %	-----	
2	All results	>>> 20	Positive	100 %
----- Total per sample: n = 20		----- Success: 100 %	-----	
Total success: 100 %				

(1036) Antigen Leb

1	All results	>>> 19	Positive	100 %
----- Total per sample: n = 19		----- Success: 100 %	-----	
2	All results	>>> 18	Negative	95 %
		1	Positive	5,3 %
----- Total per sample: n = 19		----- Success: 95 %	-----	
Total success: 95 %				

st_kl

End of report

Printed: 30.11.2017