

## EQA round: SED2/17 - Erythrocyte Sedimentation Rate

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability											
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>
<b>(541) Sedimentation /1h</b>					86						86	68	79%
Samples and groups	[mm/1hour]												
<b>Sample A</b>		65,5	17	26	86	CVP	65,5	4,5 asym.	32	99	86	81	94%
All results		65,5	17	26	86	2					86		
<b>Sample B</b>		37,9	16	43	86	CVP	37,9	4,4 asym.	18	57	86	68	79%
All results		37,9	16	43	86	0					86		
<b>(542) Sedimentation /2h</b>					85						85	70	82%
Samples and groups	[mm/2hours]												
<b>Sample A</b>		97,9	18	18	85	CVP	97,9	4,7 asym.	48	147	85	80	94%
All results		97,9	18	18	85	0					85		
<b>Sample B</b>		67,5	23	34	85	CVP	67,5	6,1 asym.	33	102	85	71	84%
All results		67,5	23	34	85	0					85		
st_kn_p					End of report								Printed: 22.09.2017