

Key: >>> ... marks correct (expected) result

> ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
--------	-------	---	--------	------------------

** ****(451) AMA**

A	(49) BioVendor	1	Positive	1,9 %
	(52) Dynex	>>> 1	Negative	1,9 %
	(87) Euroimmun	>>> 21	Negative	40 %
		1	Positive	1,9 %
	(115) BioSystems	>>> 7	Negative	13 %
	(135) OrgenTec	>>> 1	Negative	1,9 %
	(142) Immco Diagnostics	>>> 1	Negative	1,9 %
	(148) Aesku.lab Diagnostika	>>> 1	Negative	1,9 %
	(211) INOVA Diagnostics	>>> 15	Negative	29 %
	(998) home-made	>>> 1	Negative	1,9 %
	(999) another manufacturer	>>> 2	Negative	3,8 %
----- Total per sample: n = 52 -----		Success: 96 %	-----	
B	(49) BioVendor	>>> 1	Positive	1,9 %
	(52) Dynex	>>> 1	Positive	1,9 %
	(87) Euroimmun	1	Negative	1,9 %
		>>> 21	Positive	40 %
	(115) BioSystems	>>> 7	Positive	13 %
	(135) OrgenTec	>>> 1	Positive	1,9 %
	(142) Immco Diagnostics	>>> 1	Positive	1,9 %
	(148) Aesku.lab Diagnostika	>>> 1	Positive	1,9 %
	(211) INOVA Diagnostics	>>> 15	Positive	29 %
	(998) home-made	>>> 1	Positive	1,9 %
	(999) another manufacturer	>>> 2	Positive	3,8 %
----- Total per sample: n = 52 -----		Success: 98 %	-----	
Total success: 96 %				

(452) anti-LKM

A	(52) Dynex	>>> 1	Positive	2,1 %
	(87) Euroimmun	1	Negative	2,1 %
		>>> 16	Positive	34 %
	(115) BioSystems	>>> 6	Positive	13 %
	(135) OrgenTec	>>> 1	Positive	2,1 %
	(148) Aesku.lab Diagnostika	>>> 1	Positive	2,1 %
	(191) D-tek	>>> 3	Positive	6,4 %
	(211) INOVA Diagnostics	>>> 15	Positive	32 %
	(998) home-made	>>> 1	Positive	2,1 %
	(999) another manufacturer	>>> 2	Positive	4,3 %
----- Total per sample: n = 47 -----		Success: 98 %	-----	
B	(52) Dynex	>>> 1	Negative	2,1 %
	(87) Euroimmun	>>> 17	Negative	36 %
	(115) BioSystems	>>> 6	Negative	13 %
	(135) OrgenTec	>>> 1	Negative	2,1 %
	(148) Aesku.lab Diagnostika	>>> 1	Negative	2,1 %
	(191) D-tek	>>> 3	Negative	6,4 %
	(211) INOVA Diagnostics	>>> 15	Negative	32 %
	(998) home-made	>>> 1	Negative	2,1 %
	(999) another manufacturer	>>> 2	Negative	4,3 %
----- Total per sample: n = 47 -----		Success: 100 %	-----	
Total success: 98 %				

(453) SMA

A	(52) Dynex	>>> 1	Negative	1,9 %
	(87) Euroimmun	>>> 21	Negative	39 %
	(115) BioSystems	>>> 7	Negative	13 %
	(135) OrgenTec	>>> 1	Negative	1,9 %
	(148) Aesku.lab Diagnostika	>>> 1	Negative	1,9 %
	(191) D-tek	>>> 3	Negative	5,6 %
	(211) INOVA Diagnostics	>>> 16	Negative	30 %
	(998) home-made	>>> 1	Negative	1,9 %
	(999) another manufacturer	>>> 3	Negative	5,6 %
----- Total per sample: n = 54 -----		Success: 100 %	-----	
B	(52) Dynex	>>> 1	Negative	1,9 %
	(87) Euroimmun	>>> 21	Negative	39 %
	(115) BioSystems	>>> 7	Negative	13 %
	(135) OrgenTec	>>> 1	Negative	1,9 %

Test						
Sample	Group		N	Result	N _{rel}	
B	(148) Aesku.lab Diagnostika	>>>	1	Negative	1,9 %	
	(191) D-tek	>>>	3	Negative	5,6 %	
	(211) INOVA Diagnostics	>>>	16	Negative	30 %	
	(998) home-made	>>>	1	Negative	1,9 %	
	(999) another manufacturer	>>>	3	Negative	5,6 %	
	Total per sample: n = 54		Success: 100 %			
Total success: 100 %						
(455) anti-LKM1						
A	(49) BioVendor	>>>	1	Positive	1,9 %	
	(52) Dynex	>>>	1	Positive	1,9 %	
	(66) DRG	>>>	1	Positive	1,9 %	
	(87) Euroimmun	>>>	23	Positive	43 %	
	(135) OrgenTec	>>>	9	Positive	17 %	
	(148) Aesku.lab Diagnostika	>>>	3	Positive	5,6 %	
	(191) D-tek	>>>	9	Positive	17 %	
	(999) another manufacturer	>>>	7	Positive	13 %	
	Total per sample: n = 54		Success: 100 %			
	B	(49) BioVendor	>>>	1	Negative	1,9 %
(52) Dynex		>>>	1	Negative	1,9 %	
(66) DRG		>>>	1	Negative	1,9 %	
(87) Euroimmun		>>>	22	Negative	41 %	
				1	Inconclusive result (borderline)	1,9 %
(135) OrgenTec		>>>	9	Negative	17 %	
(148) Aesku.lab Diagnostika		>>>	3	Negative	5,6 %	
(191) D-tek		>>>	9	Negative	17 %	
(999) another manufacturer		>>>	7	Negative	13 %	
Total per sample: n = 54		Success: 98 %				
Total success: 98 %						
(456) AMA-M2						
A	(49) BioVendor	>>>	1	Negative	1,8 %	
	(52) Dynex	>>>	1	Negative	1,8 %	
	(66) DRG	>>>	1	Negative	1,8 %	
	(87) Euroimmun	>>>	21	Negative	38 %	
				1	Positive	1,8 %
	(101) Hycor	>>>	1	Negative	1,8 %	
	(135) OrgenTec	>>>	10	Negative	18 %	
	(148) Aesku.lab Diagnostika	>>>	1	Negative	1,8 %	
	(191) D-tek	>>>	11	Negative	20 %	
	(999) another manufacturer	>>>	7	Negative	13 %	
	Total per sample: n = 55		Success: 98 %			
B	(49) BioVendor	>>>	1	Positive	1,8 %	
	(52) Dynex	>>>	1	Positive	1,8 %	
	(66) DRG	>>>	1	Positive	1,8 %	
	(87) Euroimmun	>>>	22	Positive	40 %	
	(101) Hycor	>>>	1	Positive	1,8 %	
	(135) OrgenTec	>>>	10	Positive	18 %	
	(148) Aesku.lab Diagnostika	>>>	1	Positive	1,8 %	
	(191) D-tek	>>>	11	Positive	20 %	
	(999) another manufacturer	>>>	7	Positive	13 %	
	Total per sample: n = 55		Success: 100 %			
Total success: 98 %						
(454) anti-gp210						
A	(49) BioVendor	>>>	2	Negative	5,4 %	
	(66) DRG	>>>	1	Negative	2,7 %	
	(87) Euroimmun	>>>	19	Negative	51 %	
	(135) OrgenTec	>>>	6	Negative	16 %	
	(191) D-tek	>>>	5	Negative	14 %	
	(999) another manufacturer	>>>	4	Negative	11 %	
Total per sample: n = 37		Success: 100 %				
B	(49) BioVendor	>>>	2	Positive	5,4 %	
	(66) DRG	>>>	1	Positive	2,7 %	
	(87) Euroimmun	>>>	1	Negative	2,7 %	
				18	Positive	49 %
	(135) OrgenTec	>>>	1	Negative	2,7 %	
				5	Positive	14 %
	(191) D-tek	>>>	5	Positive	14 %	
(999) another manufacturer	>>>	4	Positive	11 %		
Total per sample: n = 37		Success: 95 %				

Test

Sample	Group		N	Result	N _{rel}
Total success: 95 %					
(457) anti-LC-1					
A	(49) BioVendor	>>>	2	Negative	4,3 %
	(52) Dynex	>>>	1	Negative	2,1 %
	(66) DRG	>>>	1	Negative	2,1 %
	(87) Euroimmun	>>>	21	Negative	45 %
	(135) OrgenTec	>>>	5	Negative	11 %
	(148) Aesku.lab Diagnostika	>>>	1	Negative	2,1 %
	(191) D-tek	>>>	11	Negative	23 %
	(999) another manufacturer	>>>	5	Negative	11 %
Total per sample: n = 47		Success: 100 %			
B	(49) BioVendor	>>>	2	Negative	4,3 %
	(52) Dynex	>>>	1	Negative	2,1 %
	(66) DRG	>>>	1	Negative	2,1 %
	(87) Euroimmun	>>>	21	Negative	45 %
	(135) OrgenTec	>>>	5	Negative	11 %
	(148) Aesku.lab Diagnostika	>>>	1	Negative	2,1 %
	(191) D-tek	>>>	11	Negative	23 %
	(999) another manufacturer	>>>	5	Negative	11 %
Total per sample: n = 47		Success: 100 %			
Total success: 100 %					
(458) anti-SLA					
A	(49) BioVendor	>>>	2	Negative	4,1 %
	(52) Dynex	>>>	1	Negative	2,0 %
	(66) DRG	>>>	1	Negative	2,0 %
	(87) Euroimmun	>>>	21	Negative	43 %
	(135) OrgenTec	>>>	6	Negative	12 %
	(147) AIDA	>>>	1	Negative	2,0 %
	(148) Aesku.lab Diagnostika	>>>	1	Negative	2,0 %
	(191) D-tek	>>>	11	Negative	22 %
	(999) another manufacturer	>>>	5	Negative	10 %
Total per sample: n = 49		Success: 100 %			
B	(49) BioVendor	>>>	2	Negative	4,1 %
	(52) Dynex	>>>	1	Negative	2,0 %
	(66) DRG	>>>	1	Negative	2,0 %
	(87) Euroimmun	>>>	21	Negative	43 %
	(135) OrgenTec	>>>	6	Negative	12 %
	(147) AIDA	>>>	1	Negative	2,0 %
	(148) Aesku.lab Diagnostika	>>>	1	Negative	2,0 %
	(191) D-tek	>>>	11	Negative	22 %
	(999) another manufacturer	>>>	5	Negative	10 %
Total per sample: n = 49		Success: 100 %			
Total success: 100 %					
(459) anti-sp100					
A	(49) BioVendor	>>>	2	Negative	5,4 %
	(66) DRG	>>>	1	Negative	2,7 %
	(87) Euroimmun	>>>	19	Negative	51 %
	(135) OrgenTec	>>>	6	Negative	16 %
	(191) D-tek	>>>	5	Negative	14 %
	(999) another manufacturer	>>>	4	Negative	11 %
Total per sample: n = 37		Success: 100 %			
B	(49) BioVendor	>>>	2	Negative	5,4 %
	(66) DRG	>>>	1	Negative	2,7 %
	(87) Euroimmun	>>>	19	Negative	51 %
	(135) OrgenTec	>>>	6	Negative	16 %
	(191) D-tek	>>>	5	Negative	14 %
	(999) another manufacturer	>>>	4	Negative	11 %
Total per sample: n = 37		Success: 100 %			
Total success: 100 %					