

Summary statistics - quantitative results

(Groups: measurement principle)

Filter: minimal size of groups n = 5

EQA round: ALB1/18 - Albumin in Urine

Dead line: 13.04.2018

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U_{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U_{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(369) Albumin					226							226	218	96%
Samples and groups	[mg/L]													
Sample A		36,5	2,9	7,8	226	CVP	36,5	0,46	21%	28,8	44,2	226	219	97%
(1) Immunochemical met.		36,5	2,9	7,8	225							225		
Other					1							1		
						1x0								
Sample B		66,3	3,8	5,8	226	CVP	66,3	0,63	21%	52,3	80,3	226	222	98%
(1) Immunochemical met.		66,3	3,9	5,8	225							225		
Other					1							1		
						1x0								
(470) ACR					212							212	201	95%
Samples and groups	[g/mol]													
Sample A		6,67	0,66	9,8	212	CVP	6,67	0,11	24%	5,06	8,28	212	205	97%
(0) Not specified		6,67	0,66	9,8	212							212		
Sample B		5,33	0,44	8,2	212	CVP	5,33	0,074	24%	4,05	6,61	212	204	96%
(0) Not specified		5,33	0,44	8,2	212							212		

st_kn_p

End of report

Printed: 17.04.2018