

## Summary statistics - quantitative results

(Groups: measurement principle)

Filter: Slovakia, minimal size of groups n = 5

## EQA round: ALB1/18 - Albumin in Urine

Dead line: 13.04.2018

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(369) Albumin</b>					33							33	31	94%
Samples and groups	[mg/L]													
<b>Sample A</b>		35,8	3,0	8,3	33	CVP	36,5	0,46	21%	28,8	44,2	33	31	94%
(1) Immunochemical met.		35,8	3,0	8,3	33	0						33		
<b>Sample B</b>		65,6	3,8	5,8	33	CVP	66,3	0,63	21%	52,3	80,3	33	32	97%
(1) Immunochemical met.		65,6	3,8	5,8	33	0						33		
<b>(470) ACR</b>					28							28	27	96%
Samples and groups	[g/mol]													
<b>Sample A</b>		6,70	0,80	12	28	CVP	6,67	0,11	24%	5,06	8,28	28	27	96%
(0) Not specified		6,70	0,80	12	28	0						28		
<b>Sample B</b>		5,30	0,53	10	28	CVP	5,33	0,074	24%	4,05	6,61	28	27	96%
(0) Not specified		5,30	0,53	10	28	0						28		

st\_kn\_p

End of report

Printed: 17.04.2018