

Summary statistics - quantitative results

(Groups: measurement principle)

Filter: minimal size of groups n = 5

EQA round: ALB2/18 - Albumin in Urine

Dead line: 12.10.2018

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(369) Albumin					227							227	213	94%
Samples and groups	[mg/L]													
Sample A		181	12	6,7	227	CVP	181	2,0	21%	142	220	227	219	96%
(1) Immunochemical met.		181	12	6,7	227	0						227		
Sample B		31,1	2,1	6,7	227	CVP	31,1	0,34	21%	24,5	37,7	227	217	96%
(1) Immunochemical met.		31,1	2,1	6,7	227	0						227		
(470) ACR					216							216	202	94%
Samples and groups	[g/mol]													
Sample A		10,8	0,98	9,0	216	CVP	10,8	0,16	24%	8,2	13,4	216	205	95%
(0) Not specified		10,8	0,98	9,0	216	0						216		
Sample B		4,29	0,39	9,1	216	CVP	4,29	0,065	24%	3,26	5,32	216	207	96%
(0) Not specified		4,29	0,39	9,1	216	0						216		

st_kn_p

End of report

Printed: 17.10.2018