

Key: >>> ... marks correct (expected) result

> ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

Test

Sample	Group		N	Result	N _{rel}
(461) anti-cardiolipin screening					
A	(2) EIA (ELISA)	>>>	18	Negative	95 %
	(99) Another measurement principle	>>>	1	Negative	5,3 %
Total per sample: n = 19		Success: 100 %			
B	(2) EIA (ELISA)	>>>	18	Positive	95 %
	(99) Another measurement principle	>>>	1	Positive	5,3 %
Total per sample: n = 19		Success: 100 %			
Total success: 100 %					

(463) anti-cardiolipin IgG

A	(2) EIA (ELISA)	>>>	51	Negative	91 %
	(4) LIA (CLIA, CMIA, ILMA)	>>>	3	Negative	5,4 %
	(99) Another measurement principle	>>>	2	Negative	3,6 %
Total per sample: n = 56		Success: 100 %			
B	(2) EIA (ELISA)	>>>	1	Negative	1,8 %
	(4) LIA (CLIA, CMIA, ILMA)	>>>	50	Positive	89 %
	(99) Another measurement principle	>>>	3	Positive	5,4 %
	(99) Another measurement principle	>>>	2	Positive	3,6 %
Total per sample: n = 56		Success: 98 %			
Total success: 98 %					

(464) anti-cardiolipin IgM

A	(2) EIA (ELISA)	>>>	50	Negative	91 %
	(4) LIA (CLIA, CMIA, ILMA)	>>>	3	Negative	5,5 %
	(99) Another measurement principle	>>>	2	Negative	3,6 %
Total per sample: n = 55		Success: 100 %			
B	(2) EIA (ELISA)	>>>	49	Negative	89 %
	(4) LIA (CLIA, CMIA, ILMA)	>>>	1	Positive	1,8 %
	(99) Another measurement principle	>>>	3	Negative	5,5 %
	(99) Another measurement principle	>>>	2	Negative	3,6 %
Total per sample: n = 55		Success: 98 %			
Total success: 98 %					

(465) anti-beta2-GP 1 screening

A	(2) EIA (ELISA)	>>>	13	Negative	93 %
	(99) Another measurement principle	>>>	1	Negative	7,1 %
Total per sample: n = 14		Success: 100 %			
B	(2) EIA (ELISA)	>>>	13	Positive	93 %
	(99) Another measurement principle	>>>	1	Positive	7,1 %
Total per sample: n = 14		Success: 100 %			
Total success: 100 %					

(467) anti-beta2-GP 1 IgG

A	(2) EIA (ELISA)	>>>	45	Negative	90 %
	(4) LIA (CLIA, CMIA, ILMA)	>>>	3	Negative	6,0 %
	(99) Another measurement principle	>>>	2	Negative	4,0 %
Total per sample: n = 50		Success: 100 %			
B	(2) EIA (ELISA)	>>>	45	Positive	90 %
	(4) LIA (CLIA, CMIA, ILMA)	>>>	3	Positive	6,0 %
	(99) Another measurement principle	>>>	2	Positive	4,0 %
Total per sample: n = 50		Success: 100 %			
Total success: 100 %					

(468) anti-beta2-GP 1 IgM

A	(2) EIA (ELISA)	>>>	43	Negative	90 %
	(4) LIA (CLIA, CMIA, ILMA)	>>>	3	Negative	6,3 %
	(99) Another measurement principle	>>>	2	Negative	4,2 %
Total per sample: n = 48		Success: 100 %			
B	(2) EIA (ELISA)	>>>	41	Negative	85 %
	(4) LIA (CLIA, CMIA, ILMA)	>>>	2	Inconclusive result (borderline)	4,2 %
	(99) Another measurement principle	>>>	3	Negative	6,3 %
	(99) Another measurement principle	>>>	2	Negative	4,2 %
Total per sample: n = 48		Success: 96 %			
Total success: 96 %					