

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U_{AV} = expanded uncertainty of the assigned value ($k = 2$)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U_{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(106) Cystatin C					11							11	9	82%
Samples and groups	[mg/L]													
Sample A		0,738	0,10	14	11	CVP	0,73	0,021	18%	0,598	0,862	11	9	82%
All results		0,738	0,10	14	11	0						11		
Sample B		2,00	0,21	11	11	CVP	2,03	0,045	18%	1,66	2,4	11	11	100%
All results		2,00	0,21	11	11	0						11		

st_kn_p

End of report

Printed: 18.04.2018