

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]						Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(690) WBC</b>					21							21	21	100%
Samples and groups	[·10(exp9)/L]													
<b>Sample A</b>		41,0	0,85	2,1	21	CVP	41	0,46	15%	34,8	47,2	21	21	100%
All results		41,0	0,85	2,1	21	0						21		
<b>Sample B</b>		21,4	0,60	2,8	21	CVP	21,4	0,32	15%	18,1	24,7	21	21	100%
All results		21,4	0,60	2,8	21	0						21		
<b>(691) CD 34+ (relative count)</b>					21							21	20	95%
Samples and groups	[%]													
<b>Sample A</b>		0,996	0,063	6,3	21	CVP	0,996	0,033	25%	0,747	1,25	21	20	95%
All results		0,996	0,063	6,3	21	0						21		
<b>Sample B</b>		0,211	0,028	13	21	CVP	0,211	0,015	25%	0,158	0,264	21	21	100%
All results		0,211	0,028	13	21	0						21		
<b>(692) CD 34+ (absolute count)</b>					21							21	20	95%
Samples and groups	[count/μL]													
<b>Sample A</b>		409	30	7,3	21	CVP	409	16	25%	306	512	21	20	95%
All results		409	30	7,3	21	0						21		
<b>Sample B</b>		44,9	4,8	11	21	CVP	44,9	2,5	25%	33,6	56,2	21	21	100%
All results		44,9	4,8	11	21	0						21		

st\_kn\_p

End of report

Printed: 29.04.2018