

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(690) WBC					21							21	21	100%
Samples and groups	[·10(exp9)/L]													
Sample A		23,0	1,1	4,6	21	CVP	23	0,56	15%	19,5	26,5	21	21	100%
All results		23,0	1,1	4,6	21	0						21		
Sample B		11,4	0,50	4,4	21	CVP	11,4	0,27	15%	9,69	13,2	21	21	100%
All results		11,4	0,50	4,4	21	0						21		
(691) CD 34+ (relative count)					21							21	21	100%
Samples and groups	[%]													
Sample A		0,385	0,037	9,5	21	CVP	0,385	0,020	25%	0,288	0,482	21	21	100%
All results		0,385	0,037	9,5	21	0						21		
Sample B		0,509	0,042	8,2	21	CVP	0,509	0,022	25%	0,381	0,637	21	21	100%
All results		0,509	0,042	8,2	21	0						21		
(692) CD 34+ (absolute count)					21							21	21	100%
Samples and groups	[count/μL]													
Sample A		88,2	9,6	11	21	CVP	88,2	5,2	25%	66,1	111	21	21	100%
All results		88,2	9,6	11	21	0						21		
Sample B		58,4	6,6	11	21	CVP	58,4	3,5	25%	43,8	73	21	21	100%
All results		58,4	6,6	11	21	0						21		

st_kn_p

End of report

Printed: 05.11.2018