

## Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: minimal size of groups n = 5

## EQA round: CRP4/18 - C-Reactive Protein

Dead line: 07.12.2018

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(100) C-reactive protein</b>					148							148	142	96%
Samples and groups	[mg/L]													
<b>Sample A</b>		54,3	4,2	7,7	148	CVP	54,3	0,84	24%	41,2	67,4	148	145	98%
(1) Abbott		53,7	2,2	4,0	22	0						22		
(12) Beckman Coulter		56,5	3,7	6,6	6	0						6		
(49) BioVendor		56,4	3,1	5,4	12	0						12		
(58) Beckman Coulter (Olympus)		57,9	2,5	4,4	37	0						37		
(60) Roche		51,6	3,9	7,6	46	0						46		
(178) DiaSys		53,7	2,7	5,0	5	0						5		
(179) Siemens (Bayer)		54,6	0,67	1,2	6	0						6		
Other					14	0						14		
<b>Sample B</b>		8,27	0,77	9,3	148	CVP	8,27	0,16	24%	6,28	10,3	148	143	97%
(1) Abbott		8,29	0,26	3,2	22	0						22		
(12) Beckman Coulter		8,55	0,22	2,6	6	0						6		
(49) BioVendor		7,60	0,96	13	12	0						12		
(58) Beckman Coulter (Olympus)		8,65	0,40	4,7	37	0						37		
(60) Roche		7,79	0,75	9,6	46	0						46		
(178) DiaSys		8,10	0,53	6,5	5	0						5		
(179) Siemens (Bayer)		9,00	0,00	0,00	6	0						6		
Other					14	1						14		

st\_kn\_p

End of report

Printed: 14.12.2018