

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U_{AV} = expanded uncertainty of the assigned value ($k = 2$)	

Test	[unit]						Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U_{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(130) WBC					26							26	26	100%
Samples and groups	f.10(exp9)/L													
Sample A		6,09	0,24	3,9	26	CVP	6,15	0,036	15%	5,22	7,08	26	26	100%
All results		6,09	0,24	3,9	26	0						26		
Sample B		3,30	0,12	3,6	26	CVP	3,34	0,019	18%	2,73	3,95	26	26	100%
All results		3,30	0,12	3,6	26	0						26		
(160) Neutrophiles					26							26	23	88%
Samples and groups	f.10(exp9)/L													
Sample A		3,97	0,24	6,0	26	CVP	3,99	0,028	25%	2,99	4,99	26	23	88%
All results		3,97	0,24	6,0	26	2						26		
Sample B		1,93	0,11	5,8	26	CVP	1,93	0,015	25%	1,44	2,42	26	23	88%
All results		1,93	0,11	5,8	26	2						26		
(161) Lymphocytes					26							26	22	85%
Samples and groups	f.10(exp9)/L													
Sample A		1,34	0,13	9,4	26	CVP	1,34	0,017	25%	1	1,68	26	22	85%
All results		1,34	0,13	9,4	26	2						26		
Sample B		1,06	0,069	6,5	26	CVP	1,07	,0097	25%	0,802	1,34	26	22	85%
All results		1,06	0,069	6,5	26	3						26		
(162) Monocytes					26							26	23	88%
Samples and groups	f.10(exp9)/L													
Sample A		0,650	0,12	19	26	CVP	0,655	0,013	50%	0,327	0,983	26	24	92%
All results		0,650	0,12	19	26	1						26		
Sample B		0,291	0,042	14	26	CVP	0,286	,0059	50%	0,143	0,429	26	23	88%
All results		0,291	0,042	14	26	0						26		
(163) Eosinophiles					26							26	19	73%
Samples and groups	f.10(exp9)/L													
Sample A		0,140	0,064	45	26	CVP	0,113	,0029	asym.	0	0,185	26	19	73%
All results		0,140	0,064	45	26	0						26		
Sample B		0,037	0,023	63	26	CVP	0,026	,0017	asym.	0	0,101	26	21	81%
All results		0,037	0,023	63	26	0						26		
(164) Basophiles					26							26	24	92%
Samples and groups	f.10(exp9)/L													
Sample A		0,032	0,018	57	26	CVP	0,048	,0040	asym.	0	0,185	26	24	92%
All results		0,032	0,018	57	26	0						26		
Sample B		0,014	0,008	62	26	CVP	0,021	,0024	asym.	0	0,101	26	24	92%
All results		0,014	0,008	62	26	0						26		