

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U_{AV} = expanded uncertainty of the assigned value ($k = 2$)	

Test	[unit]						Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U_{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(130) WBC					23							23	23	100%
Samples and groups	f.10(exp9)/L													
Sample A		3,54	0,19	5,4	23	CVP	3,54	0,023	18%	2,9	4,18	23	23	100%
All results		3,54	0,19	5,4	23	0						23	23	100%
Sample B		5,97	0,34	5,7	23	CVP	5,99	0,038	15%	5,09	6,89	23	23	100%
All results		5,97	0,34	5,7	23	0						23	23	100%
(160) Neutrophiles					23							23	23	100%
Samples and groups	f.10(exp9)/L													
Sample A		1,70	0,13	7,6	23	CVP	1,69	0,015	25%	1,26	2,12	23	23	100%
All results		1,70	0,13	7,6	23	0						23	23	100%
Sample B		3,89	0,29	7,5	23	CVP	3,94	0,036	25%	2,95	4,93	23	23	100%
All results		3,89	0,29	7,5	23	0						23	23	100%
(161) Lymphocytes					23							23	22	96%
Samples and groups	f.10(exp9)/L													
Sample A		1,15	0,11	9,6	23	CVP	1,14	0,014	25%	0,855	1,43	23	22	96%
All results		1,15	0,11	9,6	23	0						23	22	96%
Sample B		1,49	0,15	10	23	CVP	1,49	0,019	25%	1,11	1,87	23	22	96%
All results		1,49	0,15	10	23	0						23	22	96%
(162) Monocytes					23							23	22	96%
Samples and groups	f.10(exp9)/L													
Sample A		0,403	0,094	23	23	CVP	0,411	0,012	50%	0,205	0,617	23	23	100%
All results		0,403	0,094	23	23	0						23	23	100%
Sample B		0,369	0,082	22	23	CVP	0,353	0,010	50%	0,176	0,53	23	22	96%
All results		0,369	0,082	22	23	0						23	22	96%
(163) Eosinophiles					23							23	18	78%
Samples and groups	f.10(exp9)/L													
Sample A		0,248	0,039	16	23	CVP	0,253	,0052	50%	0,126	0,38	23	22	96%
All results		0,248	0,039	16	23	0						23	22	96%
Sample B		0,144	0,052	36	23	CVP	0,133	,0046	asym.	0	0,2	23	19	83%
All results		0,144	0,052	36	23	0						23	19	83%
(164) Basophiles					23							23	23	100%
Samples and groups	f.10(exp9)/L													
Sample A		0,031	0,017	56	23	CVP	0,038	,0026	asym.	0	0,107	23	23	100%
All results		0,031	0,017	56	23	0						23	23	100%
Sample B		0,041	0,027	65	23	CVP	0,052	,0037	asym.	0	0,18	23	23	100%
All results		0,041	0,027	65	23	0						23	23	100%