

Key: &gt;&gt;&gt; ... marks correct (expected) result

&gt; ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(261) Glucose</b>				
A	All results	>>> 314	Negative, normal, <= 2,8 mmol/L	100 %
Total per sample: n = 314		Success: 100 %		
B	All results	6	2,81 - 11 mmol/L	1,9 %
		> 135	11,1 - 28 mmol/L	43 %
		>>> 173	> 28 mmol/L	55 %
Total per sample: n = 314		Success: 98 %		
Total success: 98 %				
<b>(263) Ketone bodies</b>				
A	All results	>>> 314	Negative, <= 0,5 mmol/L	100 %
Total per sample: n = 314		Success: 100 %		
B	All results	> 104	1+, 2+	33 %
		>>> 210	3+, 4+	67 %
Total per sample: n = 314		Success: 100 %		
Total success: 100 %				
<b>(264) Total protein</b>				
A	All results	>>> 313	Negative, < 0,3 g/L	100 %
Total per sample: n = 313		Success: 100 %		
B	All results	1	Not specified	0,32 %
		3	0,3 - 0,5 g/L	0,96 %
		> 41	0,51 - 2,5 g/L	13 %
		>>> 268	> 2,5 g/L	86 %
Total per sample: n = 313		Success: 99 %		
Total success: 99 %				
<b>(266) Urobilinogen</b>				
A	All results	>>> 314	Negative, < 17 µmol/L	100 %
Total per sample: n = 314		Success: 100 %		
B	All results	> 53	1+, 2+	17 %
		>>> 261	3+, 4+	83 %
Total per sample: n = 314		Success: 100 %		
Total success: 100 %				
<b>(268) Bilirubin</b>				
A	All results	>>> 308	Negative	100 %
Total per sample: n = 308		Success: 100 %		
B	All results (without individual groups)	1	Not specified	0,32 %
		> 10	1+, 2+	3,2 %
		>>> 284	3+, 4+	92 %
	(223) Iris iChem Velocity	± 10	Negative	3,2 %
		± 3	1+, 2+	0,97 %
Total per sample: n = 308		Success (only evaluated results): 100 %		
Total success (only evaluated tests): 100 %				
<b>(270) WBC</b>				
A	All results	>>> 265	Negative	100 %
Total per sample: n = 265		Success: 100 %		
B	All results	3	Negative	1,1 %
		9	10 - 50 leu/µL	3,4 %
		> 74	51 - 250 leu/µL	28 %
		>>> 179	> 250 leu/µL	68 %
Total per sample: n = 265		Success: 95 %		
Total success: 95 %				
<b>(272) Blood</b>				
A	All results	1	Not specified	0,32 %
		>>> 311	Negative	99 %
		2	5 - 12 ery/µL	0,64 %
Total per sample: n = 314		Success: 99 %		
B	All results	2	5 - 12 ery/µL	0,64 %
		> 41	13 - 50 ery/µL	13 %
		>>> 271	> 50 ery/µL	86 %
Total per sample: n = 314		Success: 99 %		
Total success: 98 %				
<b>(273) pH</b>				
A	All results	>>> 269	5,0	86 %
		> 36	5,5	11 %
		> 8	6,0	2,5 %

## Test

Sample	Group	N	Result	N <sub>rel</sub>
A	All results	1	6,5	0,32 %
----- Total per sample: n = 314 ----- Success: 100 % -----				
B	All results	> 1	6,0	0,32 %
		> 41	6,5	13 %
		>>> 141	7,0	45 %
		> 109	7,5	35 %
		> 22	8,0	7,0 %
----- Total per sample: n = 314 ----- Success: 100 % -----				
Total success: 100 %				

**(274) Nitrite**

A	All results	>>> 268	Negative	100 %
----- Total per sample: n = 268 ----- Success: 100 % -----				
B	All results	>>> 268	Positive	100 %
----- Total per sample: n = 268 ----- Success: 100 % -----				
Total success: 100 %				

**(275) hCG**

A	All results	>>> 18	Negative	100 %
----- Total per sample: n = 18 ----- Success: 100 % -----				
B	All results	1	Not specified	5,6 %
		>>> 17	Positive	94 %
----- Total per sample: n = 18 ----- Success: 94 % -----				
Total success: 94 %				

st\_kl

End of report

Printed: 27.02.2018