

Key: &gt;&gt;&gt; ... marks correct (expected) result

&gt; ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(261) Glucose</b>				
A	All results	>>> 25	Negative, normal, <= 2,8 mmol/L	100 %
Total per sample: n = 25		Success: 100 %		
B	All results	1	2,81 - 11 mmol/L	4,0 %
		> 9	11,1 - 28 mmol/L	36 %
		>>> 15	> 28 mmol/L	60 %
Total per sample: n = 25		Success: 96 %		
Total success: 96 %				

**(263) Ketone bodies**

A	All results	>>> 25	Negative, <= 0,5 mmol/L	100 %
Total per sample: n = 25		Success: 100 %		
B	All results	> 8	1+, 2+	32 %
		>>> 17	3+, 4+	68 %
Total per sample: n = 25		Success: 100 %		
Total success: 100 %				

**(264) Total protein**

A	All results	>>> 25	Negative, < 0,3 g/L	100 %
Total per sample: n = 25		Success: 100 %		
B	All results	> 1	0,51 - 2,5 g/L	4,0 %
		>>> 24	> 2,5 g/L	96 %
Total per sample: n = 25		Success: 100 %		
Total success: 100 %				

**(266) Urobilinogen**

A	All results	>>> 25	Negative, < 17 µmol/L	100 %
Total per sample: n = 25		Success: 100 %		
B	All results	> 5	1+, 2+	20 %
		>>> 20	3+, 4+	80 %
Total per sample: n = 25		Success: 100 %		
Total success: 100 %				

**(268) Bilirubin**

A	All results	>>> 25	Negative	100 %
Total per sample: n = 25		Success: 100 %		
B	All results (without individual groups)	> 2	1+, 2+	8,0 %
		>>> 19	3+, 4+	76 %
	(223) Iris iChem Velocity	± 4	Negative	16 %
Total per sample: n = 25		Success (only evaluated results): 100 %		
Total success (only evaluated tests): 100 %				

**(270) WBC**

A	All results	>>> 24	Negative	100 %
Total per sample: n = 24		Success: 100 %		
B	All results	2	10 - 50 leu/µL	8,3 %
		> 5	51 - 250 leu/µL	21 %
		>>> 17	> 250 leu/µL	71 %
Total per sample: n = 24		Success: 92 %		
Total success: 92 %				

**(272) Blood**

A	All results	>>> 25	Negative	100 %
Total per sample: n = 25		Success: 100 %		
B	All results	> 1	13 - 50 ery/µL	4,0 %
		>>> 24	> 50 ery/µL	96 %
Total per sample: n = 25		Success: 100 %		
Total success: 100 %				

**(273) pH**

A	All results	>>> 19	5,0	76 %
		> 4	5,5	16 %
		> 2	6,0	8,0 %
Total per sample: n = 25		Success: 100 %		
B	All results	> 3	6,5	12 %
		>>> 11	7,0	44 %
		> 8	7,5	32 %
		> 3	8,0	12 %
Total per sample: n = 25		Success: 100 %		
Total success: 100 %				

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(274) Nitrite</b>				
A	All results	>>> 24	Negative	100 %
Total per sample: n = 24		Success: 100 %		
B	All results	>>> 24	Positive	100 %
Total per sample: n = 24		Success: 100 %		
Total success: 100 %				

**(275) hCG**

A	All results	>>> 4	Negative	100 %
Total per sample: n = 4		Success: 100 %		
B	All results	>>> 4	Positive	100 %
Total per sample: n = 4		Success: 100 %		
Total success: 100 %				