

Key: >>> ... marks correct (expected) result

> ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(261) Glucose				
A	All results	>>> 48	Negative, normal, <= 2,8 mmol/L	74 %
		> 17	2,81 - 11 mmol/L	26 %
Total per sample: n = 65		Success: 100 %		
B	All results	> 16	11,1 - 28 mmol/L	25 %
		>>> 49	> 28 mmol/L	75 %
Total per sample: n = 65		Success: 100 %		
Total success: 100 %				

(263) Ketone bodies

A	All results	>>> 65	Negative, <= 0,5 mmol/L	100 %
Total per sample: n = 65		Success: 100 %		
B	All results	2	Negative, <= 0,5 mmol/L	3,1 %
		>>> 48	1+, 2+	74 %
		> 15	3+, 4+	23 %
Total per sample: n = 65		Success: 97 %		
Total success: 97 %				

(264) Total protein

A	All results	>>> 60	Negative, < 0,3 g/L	92 %
		5	0,3 - 0,5 g/L	7,7 %
Total per sample: n = 65		Success: 92 %		
B	All results	1	0,3 - 0,5 g/L	1,5 %
		>>> 33	0,51 - 2,5 g/L	51 %
		>>> 31	> 2,5 g/L	48 %
Total per sample: n = 65		Success: 98 %		
Total success: 91 %				

(266) Urobilinogen

A	All results	>>> 65	Negative, < 17 µmol/L	100 %
Total per sample: n = 65		Success: 100 %		
B	All results	>>> 26	1+, 2+	40 %
		>>> 39	3+, 4+	60 %
Total per sample: n = 65		Success: 100 %		
Total success: 100 %				

(268) Bilirubin

A	All results	>>> 63	Negative	100 %
Total per sample: n = 63		Success: 100 %		
B	All results	> 4	1+, 2+	6,3 %
		>>> 59	3+, 4+	94 %
Total per sample: n = 63		Success: 100 %		
Total success: 100 %				

(270) WBC

A	All results	>>> 51	Negative	96 %
		1	10 - 50 leu/µL	1,9 %
		1	51 - 250 leu/µL	1,9 %
Total per sample: n = 53		Success: 96 %		
B	All results (without individual groups)	> 2	10 - 50 leu/µL	3,8 %
		> 17	51 - 250 leu/µL	32 %
		>>> 30	> 250 leu/µL	57 %
	(223) Iris iChem Velocity	± 3	Negative	5,7 %
		± 1	10 - 50 leu/µL	1,9 %
Total per sample: n = 53		Success (only evaluated results): 100 %		
Total success (only evaluated tests): 96 %				

(272) Blood

A	All results	>>> 63	Negative	98 %
		1	5 - 12 ery/µL	1,6 %
Total per sample: n = 64		Success: 98 %		
B	All results	> 2	13 - 50 ery/µL	3,1 %
		>>> 62	> 50 ery/µL	97 %
Total per sample: n = 64		Success: 100 %		
Total success: 98 %				

(273) pH

A	All results	1	5,0	1,6 %
		> 2	5,5	3,1 %
		> 16	6,0	25 %
		>>> 39	6,5	61 %

Test

Sample	Group	N	Result	N _{rel}
A	All results	> 6	7,0	9,4 %
----- Total per sample: n = 64 -----		Success: 98 %		
B	All results	1	5,5	1,6 %
		> 1	6,0	1,6 %
		> 11	6,5	17 %
		>>> 44	7,0	69 %
		> 7	7,5	11 %
----- Total per sample: n = 64 -----		Success: 98 %		
Total success: 97 %				

(274) Nitrite

A	All results	>>> 55	Negative	100 %
----- Total per sample: n = 55 -----		Success: 100 %		
B	All results	>>> 55	Positive	100 %
----- Total per sample: n = 55 -----		Success: 100 %		
Total success: 100 %				

(275) hCG

A	All results	>>> 7	Negative	100 %
----- Total per sample: n = 7 -----		Success: 100 %		
B	All results	>>> 7	Positive	100 %
----- Total per sample: n = 7 -----		Success: 100 %		
Total success: 100 %				

st_kl

End of report

Printed: 29.06.2018