

Key: >>> ... marks correct (expected) result
> ... marks conditionally correct (acceptable) result
± ... marks the result not evaluated

N ... number of the results
Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(261) Glucose				
A	All results	1	Negative, normal, <= 2,8 mmol/L	0,32 %
		4	2,81 - 11 mmol/L	1,3 %
		> 112	11,1 - 28 mmol/L	36 %
		>>> 197	> 28 mmol/L	63 %
----- Total per sample: n = 314 ----- Success: 98 % -----				
B	All results	>>> 313	Negative, normal, <= 2,8 mmol/L	100 %
		1	> 28 mmol/L	0,32 %
----- Total per sample: n = 314 ----- Success: 100 % -----				
Total success: 98 %				

(263) Ketone bodies

A	All results	1	Negative, <= 0,5 mmol/L	0,32 %
		> 129	1+, 2+	41 %
		>>> 185	3+, 4+	59 %
----- Total per sample: n = 315 ----- Success: 100 % -----				
B	All results	>>> 312	Negative, <= 0,5 mmol/L	99 %
		1	1+, 2+	0,32 %
		2	3+, 4+	0,63 %
----- Total per sample: n = 315 ----- Success: 99 % -----				
Total success: 99 %				

(264) Total protein

A	All results	1	Negative, < 0,3 g/L	0,32 %
		2	0,3 - 0,5 g/L	0,63 %
		> 30	0,51 - 2,5 g/L	9,5 %
		>>> 282	> 2,5 g/L	90 %
----- Total per sample: n = 315 ----- Success: 99 % -----				
B	All results	>>> 314	Negative, < 0,3 g/L	100 %
		1	> 2,5 g/L	0,32 %
----- Total per sample: n = 315 ----- Success: 100 % -----				
Total success: 99 %				

(266) Urobilinogen

A	All results	4	Negative, < 17 µmol/L	1,3 %
		> 61	1+, 2+	19 %
		>>> 250	3+, 4+	79 %
----- Total per sample: n = 315 ----- Success: 99 % -----				
B	All results	>>> 313	Negative, < 17 µmol/L	99 %
		2	3+, 4+	0,63 %
----- Total per sample: n = 315 ----- Success: 99 % -----				
Total success: 99 %				

(268) Bilirubin

A	All results (without individual groups)	1	Not specified	0,32 %
		3	Negative	0,97 %
		> 18	1+, 2+	5,8 %
		>>> 270	3+, 4+	87 %
	(223) Iris iChem Velocity	± 9	Negative	2,9 %
		± 7	1+, 2+	2,3 %
		± 1	3+, 4+	0,32 %
----- Total per sample: n = 309 ----- Success (only evaluated results): 99 % -----				
B	All results	>>> 307	Negative	99 %
		1	1+, 2+	0,32 %
		1	3+, 4+	0,32 %
----- Total per sample: n = 309 ----- Success: 99 % -----				
Total success (only evaluated tests): 98 %				

(270) WBC

A	All results	3	Negative	1,1 %
		15	10 - 50 leu/µL	5,6 %
		> 72	51 - 250 leu/µL	27 %
		>>> 180	> 250 leu/µL	67 %
----- Total per sample: n = 270 ----- Success: 93 % -----				
B	All results	>>> 268	Negative	99 %
		1	10 - 50 leu/µL	0,37 %
		1	> 250 leu/µL	0,37 %
----- Total per sample: n = 270 ----- Success: 99 % -----				
Total success: 93 %				

(272) Blood

Test

Sample	Group	N	Result	N _{rel}
A	All results	1	Negative	0,32 %
		4	5 - 12 ery/ μ L	1,3 %
		> 24	13 - 50 ery/ μ L	7,6 %
		>>> 286	> 50 ery/ μ L	91 %
		Total per sample: n = 315		Success: 98 %
B	All results	>>> 309	Negative	98 %
		3	5 - 12 ery/ μ L	0,95 %
		2	13 - 50 ery/ μ L	0,63 %
		1	> 50 ery/ μ L	0,32 %
		Total per sample: n = 315		Success: 98 %
Total success: 97 %				

(273) pH

A	All results	1	5,0	0,32 %
		> 2	6,0	0,63 %
		> 19	6,5	6,0 %
		>>> 64	7,0	20 %
		>>> 77	7,5	24 %
		>>> 146	8,0	46 %
		> 6	8,5	1,9 %
Total per sample: n = 315		Success: 100 %		
B	All results	>>> 269	5,0	85 %
		> 27	5,5	8,6 %
		> 17	6,0	5,4 %
		1	6,5	0,32 %
		1	8,0	0,32 %
		Total per sample: n = 315		Success: 99 %
Total success: 99 %				

(274) Nitrite

A	All results	1	Negative	0,37 %
		>>> 271	Positive	100 %
Total per sample: n = 272		Success: 100 %		
B	All results	1	Not specified	0,37 %
		>>> 270	Negative	99 %
		1	Positive	0,37 %
Total per sample: n = 272		Success: 99 %		
Total success: 99 %				

(275) hCG

A	All results	1	Negative	5,9 %
		>>> 16	Positive	94 %
Total per sample: n = 17		Success: 94 %		
B	All results	>>> 17	Negative	100 %
Total per sample: n = 17		Success: 100 %		
Total success: 94 %				

st_kl

End of report

Printed: 11.12.2018