

Key: &gt;&gt;&gt; ... marks correct (expected) result

&gt; ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(261) Glucose</b>				
A	All results	1	2,81 - 11 mmol/L	4,0 %
		>	11 11,1 - 28 mmol/L	44 %
		>>>	13 > 28 mmol/L	52 %
	Total per sample: n = 25	Success: 96 %		
B	All results	>>>	25 Negative, normal, <= 2,8 mmol/L	100 %
	Total per sample: n = 25	Success: 100 %		
Total success: 96 %				

**(263) Ketone bodies**

A	All results	>	8 1+, 2+	31 %
		>>>	18 3+, 4+	69 %
	Total per sample: n = 26	Success: 100 %		
B	All results	>>>	26 Negative, <= 0,5 mmol/L	100 %
	Total per sample: n = 26	Success: 100 %		
Total success: 100 %				

**(264) Total protein**

A	All results	>	2 0,51 - 2,5 g/L	7,7 %
		>>>	24 > 2,5 g/L	92 %
	Total per sample: n = 26	Success: 100 %		
B	All results	>>>	26 Negative, < 0,3 g/L	100 %
	Total per sample: n = 26	Success: 100 %		
Total success: 100 %				

**(266) Urobilinogen**

A	All results	>	5 1+, 2+	19 %
		>>>	21 3+, 4+	81 %
	Total per sample: n = 26	Success: 100 %		
B	All results	>>>	26 Negative, < 17 µmol/L	100 %
	Total per sample: n = 26	Success: 100 %		
Total success: 100 %				

**(268) Bilirubin**

A	All results (without individual groups)	>	1 1+, 2+	3,8 %
		>>>	21 3+, 4+	81 %
	(223) Iris iChem Velocity	±	1 Negative	3,8 %
		±	2 1+, 2+	7,7 %
		±	1 3+, 4+	3,8 %
	Total per sample: n = 26	Success (only evaluated results): 100 %		
B	All results	>>>	26 Negative	100 %
	Total per sample: n = 26	Success: 100 %		
Total success (only evaluated tests): 100 %				

**(270) WBC**

A	All results	1	10 - 50 leu/µL	4,0 %
		>	6 51 - 250 leu/µL	24 %
		>>>	18 > 250 leu/µL	72 %
	Total per sample: n = 25	Success: 96 %		
B	All results	>>>	25 Negative	100 %
	Total per sample: n = 25	Success: 100 %		
Total success: 96 %				

**(272) Blood**

A	All results	>	2 13 - 50 ery/µL	7,7 %
		>>>	24 > 50 ery/µL	92 %
	Total per sample: n = 26	Success: 100 %		
B	All results	>>>	26 Negative	100 %
	Total per sample: n = 26	Success: 100 %		
Total success: 100 %				

**(273) pH**

A	All results	>	1 6,5	3,8 %
		>>>	3 7,0	12 %
		>>>	7 7,5	27 %
		>>>	13 8,0	50 %
		>	2 8,5	7,7 %
	Total per sample: n = 26	Success: 100 %		
B	All results	>>>	21 5,0	81 %
		>	3 5,5	12 %

## Test

Sample	Group	N	Result	N <sub>rel</sub>
B	All results	>	2 6,0	7,7 %
Total per sample: n = 26		Success: 100 %		
Total success: 100 %				

**(274) Nitrite**

A	All results	>>>	25 Positive	100 %
Total per sample: n = 25		Success: 100 %		
B	All results		1 Not specified	4,0 %
		>>>	24 Negative	96 %
Total per sample: n = 25		Success: 96 %		
Total success: 96 %				

**(275) hCG**

A	All results	>>>	5 Positive	100 %
Total per sample: n = 5		Success: 100 %		
B	All results	>>>	5 Negative	100 %
Total per sample: n = 5		Success: 100 %		
Total success: 100 %				

st\_kl

End of report

Printed: 11.12.2018