

Key: &gt;&gt;&gt; ... marks correct (expected) result

&gt; ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

**Test**

Sample	Group	N	Result	N <sub>rel</sub>
<b>(223) Calprotectin</b>				
A	(2) EIA (ELISA)	>>> 27	Negative	52 %
		1	Inconclusive result (borderline)	1,9 %
		3	Positive	5,8 %
	(3) FIA (FEIA, TRACE)	>>> 2	Negative	3,8 %
	(4) LIA (CLIA, CMIA, ILMA)	>>> 4	Negative	7,7 %
	(6) IMT, IMN	>>> 4	Negative	7,7 %
	(7) HPLC, GC	>>> 9	Negative	17 %
	(99) Another measurement principle	>>> 2	Negative	3,8 %
----- Total per sample: n = 52 -----		Success: 92 %	-----	
B	(2) EIA (ELISA)	>>> 31	Positive	60 %
	(3) FIA (FEIA, TRACE)	>>> 2	Positive	3,8 %
	(4) LIA (CLIA, CMIA, ILMA)	>>> 4	Positive	7,7 %
	(6) IMT, IMN	>>> 4	Positive	7,7 %
	(7) HPLC, GC	2	Inconclusive result (borderline)	3,8 %
		>>> 7	Positive	13 %
	(99) Another measurement principle	>>> 2	Positive	3,8 %
----- Total per sample: n = 52 -----		Success: 96 %	-----	
Total success: 88 %				