

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: FC1/18 - Calprotectin in Stool

Dead line: 09.03.2018

Key: >>> ... marks correct (expected) result
> ... marks conditionally correct (acceptable) result
± ... marks the result not evaluated

N ... number of the results
Nrel ... relative number of the results

Test

Sample	Group	N	Result	N _{rel}
(223) Calprotectin				
A	(49) BioVendor	>>>	2 Negative	3,8 %
	(87) Euroimmun	>>>	1 Negative	1,9 %
	(135) OrgenTec	>>>	9 Negative	17 %
		>>>	1 Positive	1,9 %
	(164) DiaSorin	>>>	3 Negative	5,8 %
	(166) Thermo	>>>	2 Negative	3,8 %
	(219) Bühlmann	>>>	20 Negative	38 %
		>>>	2 Positive	3,8 %
	(999) another manufacturer	>>>	11 Negative	21 %
		>>>	1 Inconclusive result (borderline)	1,9 %
	----- Total per sample: n = 52 ----- Success: 92 % -----			
B	(49) BioVendor	>>>	1 Inconclusive result (borderline)	1,9 %
	>>>	1 Positive	1,9 %	
	(87) Euroimmun	>>>	1 Positive	1,9 %
	(135) OrgenTec	>>>	10 Positive	19 %
	(164) DiaSorin	>>>	3 Positive	5,8 %
	(166) Thermo	>>>	2 Positive	3,8 %
	(219) Bühlmann	>>>	1 Inconclusive result (borderline)	1,9 %
		>>>	21 Positive	40 %
	(999) another manufacturer	>>>	12 Positive	23 %
	----- Total per sample: n = 52 ----- Success: 96 % -----			
Total success: 88 %				

st_kl

End of report

Printed: 19.03.2018