

EQA round: FC1/18 - Calprotectin in Stool

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(223) Calprotectin					52							52	29	56%
Samples and groups	[µg/g]													
Sample A		27,9	14	50	52	CVP	27,9	4,8	50%	13,9	41,9	52	35	67%
All results		27,9	14	50	52	1						52		
Sample B		193	89	46	52	CVP	193	30	50%	96,5	290	52	38	73%
All results		193	89	46	52	1						52		
st_kn_p														

End of report

Printed: 19.03.2018