

## EQA round: FC2/18 - Calprotectin in Stool

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability						Comparability						
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(223) Calprotectin</b>					51							51	19	37%
Samples and groups	[µg/g]													
<b>Sample A</b>		427	240	56	51	CVP	427	85	50%	213	641	51	27	53%
All results		427	240	56	51	4						51		
<b>Sample B</b>		442	310	70	51	CVP	442	110	50%	221	663	51	26	51%
All results		442	310	70	51	2						51		
<b>Study</b>														
<b>(223) Calprotectin</b>					48							0		
Samples and groups	[µg/g]													
<b>Sample C</b>		356	270	77	48							0		
All results		356	270	77	48	10	CVP	356	110	50%	178	534	0	

st\_kn\_p

End of report

Printed: 01.10.2018