

Summary statistics - qualitative results

(Groups: manufacturer of kit)

EQA round: GP1/18 - Detection of Monoclonal Components

Dead line: 09.03.2018

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

Test

Sample	Group	N	Result	N _{rel}
(246) Gammopathies (typisation)				
A	(12) Beckman Coulter	>>>	1 IgG-kappa	1,1 %
	(34) Sebia	>>>	55 IgG-kappa	60 %
		>>>	2 IgG-kappa; kappa-free	2,2 %
	(49) BioVendor	>>>	1 IgG-kappa	1,1 %
	(91) Binding Site	>>>	1 IgG-kappa	1,1 %
	(100) Helena	>>>	3 IgG-kappa	3,3 %
	(213) Interlab Srl	>>>	29 IgG-kappa	32 %
Total per sample: n = 92		Success: 100 %		-
B	(12) Beckman Coulter	>>>	1 IgG-kappa	1,1 %
	(34) Sebia	>>>	56 IgG-kappa	61 %
		>>>	1 IgG-kappa; kappa-free	1,1 %
	(49) BioVendor	>>>	1 IgG-kappa	1,1 %
	(91) Binding Site	>>>	1 IgG-kappa	1,1 %
	(100) Helena	>>>	3 IgG-kappa	3,3 %
	(213) Interlab Srl	>>>	29 IgG-kappa	32 %
Total per sample: n = 92		Success: 100 %		-
Total success: 100 %				

st_kl

End of report

Printed: 22.03.2018