

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	$U_{AV}$ = expanded uncertainty of the assigned value ( $k = 2$ )	

Test	[unit]						Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	$U_{AV}$	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(247) Gammopathies (quan.)</b>					14							14	13	93%
Samples and groups	[g/L]													
<b>Sample A</b>		7,64	1,7	23	14	CVP	7,63	0,30	33%	5,11	10,2	14	13	93%
All results		7,64	1,7	23	14	0						14		
<b>Sample B</b>		6,40	1,4	23	14	CVP	6,17	0,28	33%	4,13	8,21	14	13	93%
All results		6,40	1,4	23	14	0						14		
<b>(279) Total protein (plasma)</b>					13							13	13	100%
Samples and groups	[g/L]													
<b>Sample A</b>		62,1	1,4	2,2	13	CVP	61,7	0,37	8%	56,7	66,7	13	13	100%
All results		62,1	1,4	2,2	13	0						13		
<b>Sample B</b>		62,3	1,0	1,7	13	CVP	61,7	0,37	8%	56,7	66,7	13	13	100%
All results		62,3	1,0	1,7	13	0						13		
<b>(314) FLC kappa</b>					11							11	11	100%
Samples and groups	[mg/L]													
<b>Sample A</b>		739	150	20	11	CVP	722	36	50%	361	1090	11	11	100%
All results		739	150	20	11	0						11		
<b>Sample B</b>		484	110	23	11	CVP	492	24	50%	246	738	11	11	100%
All results		484	110	23	11	0						11		
<b>(315) FLC lambda</b>					11							11	11	100%
Samples and groups	[mg/L]													
<b>Sample A</b>		10,1	1,0	10	11	CVP	9,69	0,53	50%	4,84	14,6	11	11	100%
All results		10,1	1,0	10	11	0						11		
<b>Sample B</b>		10,6	1,2	12	11	CVP	10,3	0,51	50%	5,15	15,5	11	11	100%
All results		10,6	1,2	12	11	0						11		

st\_kn\_p

End of report

Printed: 22.03.2018