

Summary statistics - qualitative results

(Groups: measurement principle)

EQA round: HSI/18 - Haemocoagulation Special

Dead line: 25.05.2018

Key: >>> ... marks correct (expected) result

N ... number of the results

> ... marks conditionally correct (acceptable) result

Nrel ... relative number of the results

± ... marks the result not evaluated

Test

Sample	Group	N	Result	N _{rel}
--------	-------	---	--------	------------------

Set of samples 3**(488) APC - resistance**

A3	(1) Coagulometer	>>>	48	Negative	77 %
			4	Inconclusive result (borderline)	6,5 %
			8	Positive	13 %
			2	Negative	3,2 %
	(2) Chromogenous substrat	>>>			
Total per sample: n = 62		Success: 81 %			

Total success: 81 %

st_kl

End of report

Printed: 05.06.2018