

## EQA round: HSI/18 - Haemocoagulation Special

Dead line: 25.05.2018

Key: &gt;&gt;&gt; ... marks correct (expected) result

N ... number of the results

&gt; ... marks conditionally correct (acceptable) result

Nrel ... relative number of the results

± ... marks the result not evaluated

## Test

Sample	Group	N	Result	N <sub>rel</sub>
--------	-------	---	--------	------------------

## Set of samples 3

## (488) APC - resistance

A3	(38) IL	>>>	11 Negative	18 %
			1 Positive	1,6 %
	(51) Technoclone	>>>	1 Negative	1,6 %
	(77) Skalab	>>>	1 Negative	1,6 %
	(94) Stago	>>>	2 Negative	3,2 %
	(102) Chromogenix	>>>	2 Negative	3,2 %
	(149) Siemens (Dade)	>>>	8 Negative	13 %
			4 Inconclusive result (borderline)	6,5 %
			7 Positive	11 %
	(159) Grifols	>>>	1 Negative	1,6 %
	(192) Pentapharm	>>>	22 Negative	35 %
	(193) HYPHEN BioMed	>>>	2 Negative	3,2 %

----- Total per sample: n = 62 ----- Success: 81 % -----

Total success: 81 %

st\_kl

End of report

Printed: 05.06.2018