

## Summary statistics - qualitative results

(Groups: manufacturer of instrument)

## EQA round: IFT1/18 - Immunophenotypisation

Dead line: 16.04.2018

Key: &gt;&gt;&gt; ... marks correct (expected) result

&gt; ... marks conditionally correct (acceptable) result

± ... marks the result not evaluated

N ... number of the results

Nrel ... relative number of the results

## Test

Sample	Group	N	Result	N <sub>rel</sub>
<b>(259) HLA-B27</b>				
A	(12) Beckman Coulter	1	Negative	1,5 %
		>	6 Inconclusive result (borderline)	9,0 %
	>>>	39 Positive	58 %	
	(88) Becton Dickinson	>	1 Inconclusive result (borderline)	1,5 %
		>>>	19 Positive	28 %
(177) MINDRAY	>>>	1 Positive	1,5 %	
----- Total per sample: n = 67 -----		Success: 99 %	-----	
B	(12) Beckman Coulter	>>>	45 Negative	67 %
			1 Inconclusive result (borderline)	1,5 %
	(88) Becton Dickinson	>>>	20 Negative	30 %
		(177) MINDRAY	>>>	1 Negative
	----- Total per sample: n = 67 -----		Success: 99 %	-----
Total success: 97 %				

st\_kl

End of report

Printed: 25.04.2018