

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
Set of samples 1														
(536) Prothrombin test (CoaguChek)					105							105	102	97%
Samples and groups														
Sample A1		2,09	0,085	4,1	105	CVP	2,09	0,020	20%	1,67	2,51	105	103	98%
All results		2,09	0,085	4,1	105	0						105		
Sample B1		4,10	0,37	9,0	105	CVP	4,1	0,088	20%	3,28	4,92	105	103	98%
All results		4,10	0,37	9,0	105	0						105		
st_kn_p						End of report								Printed: 27.06.2018