

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
Set of samples 1														
(536) Prothrombin test (CoaguChek)					112							112	106	95%
Samples and groups														
Sample A1		1,52	0,075	4,9	112	CVP	1,52	0,017	20%	1,21	1,83	112	107	96%
All results		1,52	0,075	4,9	112	0						112		
Sample B1		2,78	0,17	6,0	112	CVP	2,78	0,039	20%	2,22	3,34	112	106	95%
All results		2,78	0,17	6,0	112	0						112		
Set of samples 3														
(543) Prothrombin test (microINR)					31							31	26	84%
Samples and groups														
Sample A3		2,90	0,30	10	31	CVP	2,9	0,13	20%	2,32	3,48	31	30	97%
All results		2,90	0,30	10	31	0						31		
Sample B3		1,92	0,24	12	31	CVP	1,92	0,10	20%	1,53	2,31	31	26	84%
All results		1,92	0,24	12	31	0						31		
Set of samples 5														
(552) Prothrombin test (Xprecia)					22							22	21	95%
Samples and groups														
Sample A5		1,16	0,065	5,6	22	CVP	1,16	0,034	20%	0,928	1,4	22	21	95%
All results		1,16	0,065	5,6	22	0						22		
Sample B5		2,92	0,087	3,0	22	CVP	2,92	0,045	20%	2,33	3,51	22	21	95%
All results		2,92	0,087	3,0	22	0						22		

st_kn_p

End of report

Printed: 10.10.2018