

## Summary statistics - quantitative results

(Groups: measurement principle)

Filter: minimal size of groups n = 5

## EQA round: KD2/18 - Glycated Haemoglobin

Dead line: 01.06.2018

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability						Traceability															
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>	
<b>(220) Haemoglobin A1c</b>					89																89	89	100%
— Samples and groups —	[mmol/mol]																						
<b>Sample A</b>		58,9	2,3	3,9	89									CRV	58,3	1,0	16%	48,9	67,7	89	89	100%	
(1) HPLC, LC		58,8	2,1	3,6	79	0																79	
(4) Immunochemical meth.		59,7	4,8	8,1	10	0																10	
<b>Sample B</b>		77,4	2,8	3,6	89									CRV	77	1,0	16%	64,6	89,4	89	89	100%	
(1) HPLC, LC		77,2	2,5	3,3	79	0																79	
(4) Immunochemical meth.		80,3	4,8	6,0	10	0																10	

st\_kn\_np

End of report

Printed: 05.06.2018