

## Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: Slovakia, minimal size of groups n = 5

## EQA round: LMWH1/18 - Low Molecular Weight Heparin

Dead line: 25.05.2018

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(103) Low Molecular Weight Heparin</b>					12							12	10	83%
Samples and groups	[IU/mL]													
<b>Sample A</b>		0,247	0,054	22	12	CVP	0,246	,0085	30%	0,172	0,32	12	10	83%
(193) HYPHEN BioMed		0,250	0,074	30	8	0						8		
Other					4	0						4		
							1x 38, 3x 149							
<b>Sample B</b>		0,761	0,05	6,6	12	CVP	0,75	0,014	25%	0,562	0,938	12	12	100%
(193) HYPHEN BioMed		0,770	0,059	7,7	8	0						8		
Other					4	0						4		
							1x 38, 3x 149							

st\_kn\_p

End of report

Printed: 05.06.2018