

Key:	>>> ... marks correct (expected) result	N ... number of the results
	> ... marks conditionally correct (acceptable) result	Nrel ... relative number of the results
	± ... marks the result not evaluated	

Test	Group	N	Result	N _{rel}
(870) Sample A - composition				
	(3) Stan. chem. m. + polaris. mic.	>>>	3 Protein	27 %
			1 Non-identifiable	9,1 %
	(5) IR spectros. + polaris. mic.	>>>	7 Protein	64 %
	Total per sample: n = 11		Success: 91 %	
	Total success: 91 %			
(875) Sample B - composition				
	(3) Stan. chem. m. + polaris. mic.		1 Brushite; Apatit / Dahllit	9,1 %
		>>>	2 Brushite; Cystine	18 %
			1 Brushite; Cystine; Apatit / Dahllit	9,1 %
	(4) IR spectroscopy	>>>	1 Brushite; Cystine	9,1 %
	(5) IR spectros. + polaris. mic.	>>>	6 Brushite; Cystine	55 %
	Total per sample: n = 11		Success: 82 %	
	Total success: 82 %			
(876) Sample B - semi-quantity (1st component)				
	(0) Not specified	>	1 30 %	9,1 %
	(3) Stan. chem. m. + polaris. mic.	>	1 40 %	9,1 %
		>>>	1 50 %	9,1 %
		>	1 60 %	9,1 %
			1 80 %	9,1 %
	(4) IR spectroscopy		1 20 %	9,1 %
	(5) IR spectros. + polaris. mic.		1 20 %	9,1 %
		>	3 30 %	27 %
		>	1 40 %	9,1 %
	Total per sample: n = 11		Success: 73 %	
	Total success: 73 %			
(877) Sample B - semi-quantity (2nd component)				
	(3) Stan. chem. m. + polaris. mic.	>	1 40 %	10 %
		>>>	2 50 %	20 %
	(4) IR spectroscopy		1 80 %	10 %
	(5) IR spectros. + polaris. mic.	>	1 60 %	10 %
		>	4 70 %	40 %
			1 80 %	10 %
	Total per sample: n = 10		Success: 80 %	
	Total success: 80 %			
(880) Sample C - composition				
	(3) Stan. chem. m. + polaris. mic.	>	2 Whewellite; Uricite	18 %
		>	1 Whewellite; Monosodium urate monohydrate; Uricite	9,1 %
			1 Whewellite; Apatit / Dahllit	9,1 %
	(5) IR spectros. + polaris. mic.	>>>	7 Whewellite; Monosodium urate monohydrate	64 %
	Total per sample: n = 11		Success: 91 %	
	Total success: 91 %			
(881) Sample C - semi-quantity (1st component)				
	(3) Stan. chem. m. + polaris. mic.		1 40 %	9,1 %
			1 50 %	9,1 %
		>	1 70 %	9,1 %
		>	1 90 %	9,1 %
	(5) IR spectros. + polaris. mic.		1 20 %	9,1 %
			1 30 %	9,1 %
			1 40 %	9,1 %
		>	4 70 %	36 %
	Total per sample: n = 11		Success: 55 %	
	Total success: 55 %			
(882) Sample C - semi-quantity (2nd component)				
	(3) Stan. chem. m. + polaris. mic.	>	1 10 %	13 %
	(5) IR spectros. + polaris. mic.	>	4 30 %	50 %
			1 60 %	13 %
			1 70 %	13 %
			1 80 %	13 %
	Total per sample: n = 8		Success: 63 %	
	Total success: 63 %			
(883) Sample C - semi-quantity (3rd component)				
	(3) Stan. chem. m. + polaris. mic.	>	1 30 %	33 %

Summary statistics - qualitative results

(Groups: measurement principle)

EQA round: MK2/18 - Urinary Calculi Analysis

Dead line: 07.12.2018

Test

Group	N	Result	N _{rel}
(3) Stan. chem. m. + polaris. mic.	> 1	40 %	33 %
	1	60 %	33 %
Total per sample: n = 3		Success: 67 %	
Total success: 67 %			

st_kl

End of report

Printed: 12.12.2018