

## EQA round: RET2/18 - Reticulocytes (Microscopy)

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U <sub>AV</sub> = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	U <sub>AV</sub>	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(211) Reticulocytes</b>					82							82	75	91%
Samples and groups	[·10(exp9)/L]													
<b>Sample A</b>		136	20	14	82	CVP	136	5,3	50%	68	204	82	77	94%
All results		136	20	14	82	0						82		
<b>Sample B</b>		288	58	20	82	CVP	288	16	50%	144	432	82	80	98%
All results		288	58	20	82	0						82		
st_kn_p						End of report								Printed: 04.10.2018