

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: minimal size of groups n = 5

EQA round: RFA2/18 - Risk Factors for Atherosclerosis

Dead line: 31.08.2018

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(111) Cholesterol					285						285	272	95%	
Samples and groups	[mmol/L]													
Sample A		3,32	0,12	3,7	285	CVP	3,32	0,018	8%	3,05	3,59	285	278	98%
(1) Abbott		3,37	0,073	2,1	44	0						44		
(12) Beckman Coulter		3,36	0,089	2,6	9	0						9		
(46) Erba Lachema		3,29	0,14	4,2	13	0						13		
(49) BioVendor		3,40	0,15	4,4	7	0						7		
(58) Beckman Coulter (Olympus)		3,42	0,13	3,7	63	0						63		
(60) Roche		3,24	0,075	2,3	96	0						96		
(149) Siemens (Dade)		3,19	0,18	5,7	11	0						11		
(177) MINDRAY		3,36	0,067	2,0	7	0						7		
(179) Siemens (Bayer)		3,30	0,10	3,1	24	0						24		
Other					11	0						11		
							2x 5, 1x 18, 1x 32, 1x 121, 3x 166, 3x 178							
Sample B		4,89	0,18	3,6	285	CVP	4,89	0,025	8%	4,49	5,29	285	276	97%
(1) Abbott		4,97	0,09	1,8	44	0						44		
(12) Beckman Coulter		4,93	0,19	3,9	9	0						9		
(46) Erba Lachema		4,86	0,13	2,7	13	0						13		
(49) BioVendor		5,00	0,15	3,0	7	0						7		
(58) Beckman Coulter (Olympus)		5,04	0,16	3,2	63	0						63		
(60) Roche		4,77	0,12	2,6	96	0						96		
(149) Siemens (Dade)		4,79	0,22	4,5	11	0						11		
(177) MINDRAY		4,99	0,089	1,8	7	0						7		
(179) Siemens (Bayer)		4,84	0,15	3,1	24	0						24		
Other					11	0						11		
							2x 5, 1x 18, 1x 32, 1x 121, 3x 166, 3x 178							
(112) Triglycerides					285							285	277	97%
Samples and groups	[mmol/L]													
Sample A		1,05	0,046	4,4	285	CVP	1,05	,0067	15%	0,892	1,21	285	280	98%
(1) Abbott		1,02	0,041	4,0	43	0						43		
(12) Beckman Coulter		1,01	0,03	2,9	8	0						8		
(46) Erba Lachema		1,07	0,047	4,4	16	0						16		
(49) BioVendor		1,00	0,03	3,0	7	0						7		
(58) Beckman Coulter (Olympus)		1,04	0,033	3,2	60	0						60		
(60) Roche		1,08	0,03	2,8	96	0						96		
(149) Siemens (Dade)		1,05	0,10	9,6	11	0						11		
(177) MINDRAY		1,04	0,022	2,1	7	0						7		
(179) Siemens (Bayer)		1,06	0,021	2,0	24	0						24		
Other					13	0						13		
							3x 5, 1x 18, 2x 32, 1x 121, 3x 166, 3x 178							
Sample B		1,53	0,056	3,7	285	CVP	1,53	,0081	15%	1,3	1,76	285	278	98%
(1) Abbott		1,50	0,05	3,3	43	0						43		
(12) Beckman Coulter		1,51	0,082	5,4	8	0						8		
(46) Erba Lachema		1,55	0,062	4,0	16	0						16		
(49) BioVendor		1,47	0,052	3,5	7	0						7		
(58) Beckman Coulter (Olympus)		1,51	0,039	2,6	60	0						60		
(60) Roche		1,56	0,039	2,5	96	0						96		
(149) Siemens (Dade)		1,54	0,10	6,7	11	0						11		
(177) MINDRAY		1,51	0,052	3,4	7	0						7		
(179) Siemens (Bayer)		1,54	0,045	2,9	24	0						24		
Other					13	0						13		
							3x 5, 1x 18, 2x 32, 1x 121, 3x 166, 3x 178							
(113) Cholesterol HDL					287							287	280	98%
Samples and groups	[mmol/L]													
Sample A		0,881	0,046	5,2	287	CVP	0,881	,0066	15%	0,748	1,02	287	281	98%
(1) Abbott		0,877	0,041	4,7	41	0						41		
(12) Beckman Coulter		0,895	0,052	5,8	8	0						8		
(46) Erba Lachema		0,901	0,018	2,0	16	0						16		
(49) BioVendor		0,920	0,03	3,2	7	0						7		
(58) Beckman Coulter (Olympus)		0,845	0,039	4,6	60	0						60		
(60) Roche		0,894	0,027	3,0	94	0						94		
(149) Siemens (Dade)		0,957	0,034	3,5	11	0						11		
(177) MINDRAY		0,905	0,022	2,5	6	0						6		
(179) Siemens (Bayer)		0,838	0,03	3,6	24	0						24		
Other					20	0						20		
							2x 5, 1x 18, 1x 32, 4x 70, 1x 75, 1x 121, 3x 158, 2x 166, 4x 178, 1x 999							
Sample B		1,30	0,062	4,8	287	CVP	1,3	,0090	15%	1,1	1,5	287	280	98%

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: minimal size of groups n = 5

EQA round: RFA2/18 - Risk Factors for Atherosclerosis

Dead line: 31.08.2018

Test	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability								
							AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}	
(113) Cholesterol HDL					287							287	280	98%	
Samples and groups	[mmol/L]														
Sample B		1,30	0,062	4,8	287		CVP	1,3	0,090	15%	1,1	1,5	287	280	98%
(1) Abbott		1,30	0,063	4,8	41	0							41		
(12) Beckman Coulter		1,36	0,096	7,1	8	0							8		
(46) Erba Lachema		1,35	0,033	2,4	16	0							16		
(49) BioVendor		1,35	0,045	3,3	7	0							7		
(58) Beckman Coulter (Olympus)		1,25	0,056	4,5	60	0							60		
(60) Roche		1,31	0,045	3,4	94	0							94		
(149) Siemens (Dade)		1,36	0,048	3,5	11	0							11		
(177) MINDRAY		1,34	0,022	1,7	6	0							6		
(179) Siemens (Bayer)		1,30	0,038	2,9	24	0							24		
Other					20	0							20		
															2x 5, 1x 18, 1x 32, 4x 70, 1x 75, 1x 121, 3x 158, 2x 166, 4x 178, 1x 999
(114) Apolipoprotein AI					84								84	82	98%
Samples and groups	[g/L]														
Sample A		1,05	0,075	7,1	84		CVP	1,05	0,020	21%	0,829	1,28	84	82	98%
(1) Abbott		1,08	0,03	2,8	8	0							8		
(12) Beckman Coulter		1,07	0,10	9,7	7	0							7		
(58) Beckman Coulter (Olympus)		1,06	0,11	11	17	0							17		
(60) Roche		1,02	0,051	5,0	29	0							29		
(149) Siemens (Dade)		1,00	0,10	10	5	0							5		
(179) Siemens (Bayer)		1,07	0,13	12	9	0							9		
Other					9	0							9		
															1x 5, 2x 46, 3x 49, 1x 116, 1x 177, 1x 178
Sample B		1,50	0,11	7,0	84		CVP	1,5	0,028	21%	1,18	1,82	84	84	100%
(1) Abbott		1,51	0,096	6,4	8	0							8		
(12) Beckman Coulter		1,50	0,14	9,4	7	0							7		
(58) Beckman Coulter (Olympus)		1,41	0,12	8,5	17	0							17		
(60) Roche		1,51	0,062	4,1	29	0							29		
(149) Siemens (Dade)		1,44	0,089	6,2	5	0							5		
(179) Siemens (Bayer)		1,50	0,15	9,9	9	0							9		
Other					9	0							9		
															1x 5, 2x 46, 3x 49, 1x 116, 1x 177, 1x 178
(115) Apolipoprotein B					99								99	95	96%
Samples and groups	[g/L]														
Sample A		0,668	0,038	5,6	99		CVP	0,668	0,093	18%	0,547	0,789	99	96	97%
(1) Abbott		0,692	0,04	5,7	11	0							11		
(12) Beckman Coulter		0,695	0,052	7,5	8	0							8		
(58) Beckman Coulter (Olympus)		0,652	0,057	8,7	20	0							20		
(60) Roche		0,660	0,016	2,5	36	0							36		
(149) Siemens (Dade)		0,635	0,052	8,2	5	0							5		
(179) Siemens (Bayer)		0,700	0,03	4,2	9	0							9		
Other					10	0							10		
															1x 5, 2x 46, 3x 49, 1x 116, 1x 177, 2x 178
Sample B		0,959	0,049	5,1	99		CVP	0,959	0,012	18%	0,786	1,14	99	96	97%
(1) Abbott		0,977	0,045	4,6	11	0							11		
(12) Beckman Coulter		1,00	0,03	3,0	8	0							8		
(58) Beckman Coulter (Olympus)		0,895	0,058	6,4	20	0							20		
(60) Roche		0,971	0,027	2,7	36	0							36		
(149) Siemens (Dade)		0,920	0,089	9,7	5	0							5		
(179) Siemens (Bayer)		0,980	0,045	4,5	9	0							9		
Other					10	0							10		
															1x 5, 2x 46, 3x 49, 1x 116, 1x 177, 2x 178
(116) Cholesterol LDL (calculation)					139								139	131	94%
Samples and groups	[mmol/L]														
Sample A		1,97	0,14	7,1	139		CVP	1,97	0,029	15%	1,67	2,27	139	133	96%
(0) not specified		1,97	0,14	7,1	139	1							139		
Sample B		2,90	0,19	6,6	139		CVP	2,9	0,040	15%	2,46	3,34	139	134	96%
(0) not specified		2,90	0,19	6,6	139	1							139		
(118) Cholesterol LDL (direct determ.)					198								198	194	98%
Samples and groups	[mmol/L]														
Sample A		2,02	0,10	4,9	198		CVP	2,02	0,017	15%	1,71	2,33	198	195	98%
(1) Abbott		2,03	0,048	2,4	25	0							25		
(46) Erba Lachema		2,05	0,096	4,7	12	0							12		
(49) BioVendor		1,90	0,015	0,78	6	0							6		
(58) Beckman Coulter (Olympus)		2,11	0,071	3,4	50	0							50		
(60) Roche		1,95	0,061	3,1	58	0							58		
(149) Siemens (Dade)		2,03	0,11	5,5	7	0							7		
(179) Siemens (Bayer)		1,97	0,086	4,4	15	0							15		
Other					25	0							25		
															2x 5, 4x 12, 3x 32, 2x 70, 4x 158, 1x 166, 4x 177, 3x 178, 1x 196, 1x 999
Sample B		2,97	0,16	5,4	198		CVP	2,97	0,028	15%	2,52	3,42	198	194	98%

Summary statistics - quantitative results

(Groups: manufacturer of kit)

Filter: minimal size of groups n = 5

EQA round: RFA2/18 - Risk Factors for Atherosclerosis

Dead line: 31.08.2018

Test	[unit]	RoM	SD	CV [%]	N _{tot}	N _{out}	Comparability								
							AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}	
(118) Cholesterol LDL (direct determ.)	[mmol/L]				198							198	194	98%	
----- Samples and groups -----															
Sample B		2,97	0,16	5,4	198		CVP	2,97	0,028	15%	2,52	3,42	198	194	98%
(1) Abbott		3,01	0,092	3,1	25	0							25		
(46) Erba Lachema		3,03	0,19	6,2	12	0							12		
(49) BioVendor		2,78	0,045	1,6	6	0							6		
(58) Beckman Coulter (Olympus)		3,11	0,11	3,6	50	0							50		
(60) Roche		2,87	0,095	3,3	58	0							58		
(149) Siemens (Dade)		2,92	0,13	4,3	7	0							7		
(179) Siemens (Bayer)		2,93	0,17	5,6	15	0							15		
Other					25	0							25		
----- 2x 5, 4x 12, 3x 32, 2x 70, 4x 158, 1x 166, 4x 177, 3x 178, 1x 196, 1x 999 -----															
(117) Lipoprotein (a) [g/L]	[g/L]				25								25	25	100%
----- Samples and groups -----															
Sample A		0,185	0,014	7,8	25		CVP	0,185	,0070	32%	0,125	0,245	25	25	100%
(116) Sentinel Diagnostics		0,185	0,008	4,4	5	0							5		
(179) Siemens (Bayer)		0,194	0,006	3,5	5	0							5		
Other					15	0							15		
----- 3x 12, 2x 58, 2x 60, 4x 75, 2x 91, 1x 152, 1x 178 -----															
Sample B		0,266	0,019	7,2	25		CVP	0,266	,0094	32%	0,18	0,352	25	25	100%
(116) Sentinel Diagnostics		0,260	0,015	5,7	5	0							5		
(179) Siemens (Bayer)		0,265	0,007	2,8	5	0							5		
Other					15	0							15		
----- 3x 12, 2x 58, 2x 60, 4x 75, 2x 91, 1x 152, 1x 178 -----															
(119) Lipoprotein (a) [nmol/L]	[nmol/L]				23								23	21	91%
----- Samples and groups -----															
Sample A		33,7	2,7	8,0	23		CVP	33,7	1,4	20%	26,9	40,5	23	21	91%
(60) Roche		33,8	2,2	6,4	19	0							19		
Other					4	0							4		
----- 1x 1, 1x 75, 2x 116 -----															
Sample B		51,7	3,7	7,2	23		CVP	51,7	1,9	20%	41,3	62,1	23	21	91%
(60) Roche		51,8	2,9	5,5	19	0							19		
Other					4	0							4		
----- 1x 1, 1x 75, 2x 116 -----															

st_kn_p

End of report

Printed: 13.09.2018