

## Summary statistics - quantitative results

(Groups: measurement principle)

Filter: Slovakia, minimal size of groups n = 5

## EQA round: ALB1/19 - Albumin in Urine

Dead line: 12.04.2019

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	$U_{AV}$ = expanded uncertainty of the assigned value ( $k = 2$ )	

Test	[unit]	Comparability					Comparability							
		RoM	SD	CV [%]	N <sub>tot</sub>	N <sub>out</sub>	AV	$U_{AV}$	D <sub>max</sub>	LL	UL	N <sub>eva</sub>	N <sub>suc</sub>	S <sub>rel</sub>
<b>(369) Albumin</b>					39							39	35	90%
Samples and groups	[mg/L]													
<b>Sample A</b>		59,5	4,8	8,1	39	CVP	59	0,55	21%	46,6	71,4	39	37	95%
(1) Immunochemical met.		59,6	5,2	8,7	36							36		
Other					3							3		
						3x99								
<b>Sample B</b>		15,0	2,1	14	39	CVP	14,7	0,26	30%	10,2	19,2	39	36	92%
(1) Immunochemical met.		15,0	2,3	15	36							36		
Other					3							3		
						3x99								
<b>(470) ACR</b>					34							34	30	88%
Samples and groups	[g/mol]													
<b>Sample A</b>		4,66	0,49	11	34	CVP	4,62	0,055	24%	3,51	5,73	34	32	94%
(0) Not specified		4,66	0,49	11	34							34		
<b>Sample B</b>		2,60	0,40	15	34	CVP	2,54	0,055	30%	1,77	3,31	34	30	88%
(0) Not specified		2,60	0,40	15	34							34		

st\_kn\_p

End of report

Printed: 18.04.2019