

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]						Comparability							
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(690) WBC					21							21	21	100%
Samples and groups	[·10(exp9)/L]													
Sample A		18,5	0,58	3,1	21	CVP	18,5	0,31	15%	15,7	21,3	21	21	100%
All results		18,5	0,58	3,1	21	0						21		
Sample B		23,2	0,75	3,2	21	CVP	23,2	0,40	15%	19,7	26,7	21	21	100%
All results		23,2	0,75	3,2	21	0						21		
(691) CD 34+ (relative count)					21							21	21	100%
Samples and groups	[%]													
Sample A		0,836	0,061	7,2	21	CVP	0,836	0,032	25%	0,627	1,05	21	21	100%
All results		0,836	0,061	7,2	21	0						21		
Sample B		0,536	0,031	5,8	21	CVP	0,536	0,017	25%	0,402	0,67	21	21	100%
All results		0,536	0,031	5,8	21	0						21		
(692) CD 34+ (absolute count)					21							21	21	100%
Samples and groups	[count/μL]													
Sample A		155	13	8,4	21	CVP	155	6,9	25%	116	194	21	21	100%
All results		155	13	8,4	21	0						21		
Sample B		125	7,2	5,7	21	CVP	125	3,8	25%	93,7	157	21	21	100%
All results		125	7,2	5,7	21	0						21		

st_kn_p

End of report

Printed: 18.04.2019