

Summary statistics - quantitative results

(Groups: measurement principle)

Filter: minimal size of groups n = 5

EQA round: CRP2/19 - C-Reactive Protein

Dead line: 31.05.2019

RoM = robust average	AV = assigned value	Dmax = acceptable percent difference
SD = standard deviation	CRV = certified reference value	LL = lower limit
CV = coefficient of variation	RV = reference value	UL = upper limit
Ntot = total number of participants	CVE = consensus value from experts	Neva = number of evaluated participants
Nout = number of results excluded before calculation	CVP = consensus value from all participants	Nsuc = number of successful participants
	CVPG = consensus value from participants groups	Srel = success (relative)
	U _{AV} = expanded uncertainty of the assigned value (k = 2)	

Test	[unit]	Comparability												
		RoM	SD	CV [%]	N _{tot}	N _{out}	AV	U _{AV}	D _{max}	LL	UL	N _{eva}	N _{suc}	S _{rel}
(100) C-reactive protein					144							144	141	98%
Samples and groups	[mg/L]													
Sample A		9,27	0,81	8,8	144	CVP	9,27	0,17	24%	7,04	11,5	144	141	98%
(2) Immunoturbidimetry		9,29	0,82	8,8	137							137		
Other					7							7		
-----						4x 3, 3x 99								
Sample B		20,6	1,6	7,7	144	CVP	20,6	0,32	24%	15,6	25,6	144	142	99%
(2) Immunoturbidimetry		20,7	1,6	7,6	137							137		
Other					7							7		
-----						4x 3, 3x 99								
st_kn_p						End of report								Printed: 04.06.2019